

AMENDMENT 1

PURCHASE OF ONLINE GAS ANALYZER

Ref. No. CHE/16-17/144/MUAY/RVIN

Due date: 3 Oct. 2017

The revised gas composition is as follows:

Gas	Concentration range	Resolution/Accuracy
O ₂	0-100%	0.1% / ±0.5-1%
CO	0-50%	0.1% / ±0.5-1%
CO ₂	0-50%	0.1% / ±0.5-1%
CH ₄	0-20%	0.1% / ±0.5-1%
H ₂	0-50%	0.1% / ±0.5-1%

All other technical specifications and commercial terms remain unaltered.

Dr. R. Vinu

Associate Professor

Chemical Engineering

IIT Madras

Ph: +91-44-22574187

E-mail: vinu@iitm.ac.in