

Tender specifications of the He-Ne Laser Tube and accessories

Specifications		Compliance (Yes/No)	Reference Page No.
He-Ne laser tube	He-Ne laser tube with Brewster window termination		
Wavelength	632.8 nm		
Electrical power supply	Power supply for above tubes		
Output Power level	5.0 mW		
Accessories (must be provided separately)	<ol style="list-style-type: none"> 1. Alden connector cable, 2. Ballast resistor 3. output coupler and high reflector 		
The equipment should run with warm up period of 15min <ol style="list-style-type: none"> 1. At room temperature (23-30 C) 2. With regular power supply of 220-230V AC 			
Other Conditions			
1. Installation and local support must be provided.			
2. Should provide 1- year warranty.			

(Note: It is mandatory for the bidders to provide the compliance statement in tabular column format along with catalogue page number (comply/not comply) for the Above points with document proof as required. Failing which bidders will be technically disqualified)