

## MAIN COMPONENTS

The major equipment comprise of the following:

Si. No.	Main Component/Equipment	Unit	Quantity
1.	Endoscopy arm along with accessories	No.	1

## TECHNICAL SPECIFICATIONS FOR Endoscopy arm along with accessories

Feature	Required features
Application	To delivery laser light from source to test section with smooth movements without loss of light intensity
Type of Endoscopy	arm
Number of joints	7
Clear aperture	19 mm
Mirrors – diameter x thickness	Diameter : 28 mm Thickness : 3 mm
Adjustment tolerance – position	< 1 mm
Adjustment tolerance – angle	< 1 mrad
Laser light source	Double pulsed laser
Warranty	Minimum one year warranty period from the date of installation.
<b>Eligibility Criteria</b>	a.Should have Supplied to any IIT in India & details to be provided. b. Reputed Make

\*\* - Bidder has to specify in their offer

**Note:** All the technical parameters are indicative value, Bidder will be designing and specifying in their offer.