

Open Floating Tender-Technical Specification for Accelerometer:

Accelerometer total 5 No's:

1. ICB or PCB Type it should reach standard Water Proofing Terminals.
2. Analog output Range min range of 1V/g / max 10V (No Current Output).
3. Real time Output Frequency Response 0.06 to 450 HZ.
4. Standard BNC out Put connectors.
5. Strictly Miniature Type / with standard fixing accessories (Low noise connecting cable / connecting screws, etc..../).

Quote should be provided separately for under water and without under water in separate envelope.

Place: IIT Madras

Date: 04/06/2019.

Reference Name: Dr. Srirani.

Contact No : 044-22574813

