

	<p style="text-align: center;"><b>INDIAN INSTITUTE OF TECHNOLOGY MADRAS</b> Chennai 600 036</p> <p style="text-align: center;">Telephone: [044] 2257 4274/9798/23 E-mail: <a href="mailto:tender@imail.iitm.ac.in">tender@imail.iitm.ac.in</a> / <a href="mailto:arpp@iitm.ac.in">arpp@iitm.ac.in</a></p>	 <p style="text-align: center;"><b>RWTUV</b></p>
---	---	---

22.05.2019

## Department of Chemistry

### Addendum-1

**Tender Reference no:** CHY/BRAJ/006/2019


**Name of the Item:** The Dye laser for CRD spectrometer

#### **Addendum details:**

Amplification in the dye laser should preferably be done using Bethune cell for a best beam profile, which is very critical and essential for our experiments.

All other terms and conditions remain the same.

**Tender Inviting Authority:**



Senior Manager (Project Purchase)  
IC&SR, 2<sup>nd</sup> Floor, IIT Madras  
Chennai 600036