

Technical Specifications for fabrication of tank – 1 Unit**1.0 Bidder Eligibility Criteria-I**

Si.No	Bidder Eligibility Criteria-I	Complied/Not Complied	Ref pg.No
1.	The bidder/OEM should have supplied a similar items to IITs, NITs, IISERs, CSIR Labs or other Govt. R&D organizations in the last 5 years, PO copies or installation certificates along with contact details of end user need to be submitted as the proof of supply. IIT Madras reserves its right to verify the claims submitted by the bidder and the feedback from the previous customers will be part of technical evaluation		

2.0 Technical Specifications II

Si.No	Specification	Description	Complied/Not Complied	Ref pg.No
1.	Fabrication of tank	Drawing attached		
2.	Material	SS316L		
3.	Sheet thickness	6 mm		
4.	Minimum vacuum pressure	40 mbar		
5.	Quantity	1 Unit		