

TECHNICAL BID PROFORMA
Item Name: Walk in Environmental Chamber

1.0 Bidder Eligibility Criteria:

I	Bidder Eligibility Criteria-I (Public Procurement – Preference to Make in India)	Class I / Class II	Local Content value	Reference, Page No.
I	Only 'Class-I local suppliers' and 'Class-II local suppliers', as defined under DIPP, MoCI Order No. P-45021/2/2017-PP (BE II) dated 16 th September 2020 and other subsequent orders issued therein.			
2.0	Bidder Eligibility Criteria-II	Complied/Not Complied	Reference Page No.	Remarks, If any
1	The supplier/vendor must be an original equipment manufacturer or an authorized agent/dealer/seller of the item. The vendor should have supplied at least 5 similar units at other IITs or NITs, or national laboratories and research centres (DRDO/CSIR/BARC/IGCAR), or R&D centres of reputed multinational companies, or globally recognized universities, in the last 5 years. PO copy, or performance certificate, along with contact details (for these organizations) needs to be submitted			

3.0 Technical Compliance:

Sl.No.	Specifications	Complied/Not Complied	Reference Page No.
A. Temperature and humidity control:			
1.	Temperature range: - 20 to + 50 °C, with an accuracy of 0.5 °C		
2.	Humidity range: 20 to 95% in the temperature range of 20 to 50 °C, with an accuracy of 2%		
3.	Programmable control of temperature and humidity cycles should be possible, apart from setting up of the chamber at specific set point temperature and humidity		
4.	Alarms should be enabled in case of overshoot of temperature and / or humidity		
B. Chamber specifications:			
1.	Insulated panels for the chamber walls should be at least 100 mm thick, and modular design should be enabled for easy transport and set up		
2.	Chamber floor should be rugged and capable of supporting loads up to 30 kN/m ²		
3.	Internal size of the chamber should be at least 2 m (width) x 2.5 m (height) x 6 m (length)		
4.	Chamber door should be at least 1.5 m wide and 2 m tall		
C. Data display and recording:			
1.	(i) Touch screen display of at least 15 cm size should be provided. The display should also include ports for Ethernet, USB, Serial and SD card connections		
2.	Real time and historical colour graph displays should be enabled		

3.	System software should enable remote access through mobile phone or laptop		
4.	Warranty:	Essential: Compressor(s) – 5 years; All other parts – 3 years	
5.		Optional Quote-1: Additional 2 years warranty may be quoted separately	
6.		Optional Quote-2: AMC beyond 3 Years Warranty period	
7.		(Optional quotes will not be taken up for price bid comparison)	
Other Terms and Conditions			
1.	The system should be delivered within 14-16 weeks from the opening of the letter of credit or issue of the purchase order, whichever is later.		
2.	Costs and related information should be given only in the financial bid.		
3.	The cost should include all delivery costs up to IIT Madras.		
4.	The warranty shall commence only from the date of equipment installation at IITM.		
5.	IIT Madras reserves the right to exclude some items from the purchase.		
6.	As part of tender technical evaluation, IIT Madras will approach the past end users for feedback and in case of any adverse feedback the bidder will be technically disqualified.		
7.	The system should be installed and commissioned with no additional cost.		
8.	Training at IIT Madras should be provided with no additional cost.		
9.	System manual should be provided in CD or pen drive form.		
10.	Services and spares should be necessarily available within India.		

S.No	Technical Bid should comprise of the following	Complied/Not Complied	Reference Page No.
1.	Detailed Technical brochure		
2.	Detailed technical write up explaining how each of the Technical Specifications are complied with, indicating the location in the brochure		
3.	The vendor should guarantee round the clock technical support not only during the warranty period but even beyond through an annual maintenance contract. Demonstration of having provided such satisfactory technical support to customers shall be enclosed with the technical bid.		
4.	The manufacturer must have a well-qualified technical support team		

(Note: It is mandatory for the bidders to provide the compliance statement in tabular column format along with catalogue page number (comply/not comply) for the above points with document proof as required. Failing which bidders will be technically disqualified)

**SIGNATURE OF BIDDER ALONG WITH
SEAL OF THE COMPANY WITH DATE**