

## Technical Specifications of Water Volume Flow Meters-3 Units

### 1.0 Bidder Eligibility Criteria-I

Sl. No	Bidder Eligibility Criteria-I	Complied / Not Complied	Reference Page No.	Remarks, If any
1	The bidder/OEM should have supplied a similar items to IITs, NITs, IISERs, CSIR Labs or other Govt. R&D organizations in the last 5 years, PO copies or installation certificates along with contact details of end user need to be submitted as the proof of supply. IIT Madras reserves its right to verify the claims submitted by the bidder and the feedback from the previous customers will be part of technical evaluation			

### 2.0 Technical Specifications II

Sl. NO	Feature	Required features	Complied / Not Complied	Reference, Page. No.
1.	Working fluid	Water		
2.	Working fluid temperature range	0 to 100°C		
3.	Working fluid pressure range	1 to 8 bar		
4.	Type	Rotameter / float		
5.	Material	Glass		
6.	Density	998 kg/m <sup>3</sup>		
7.	Ambient temperature	20 to 50°C		
8.	Mass flow rate range	0.5 to 3.5 kg/s		
9.	Line size	1 inch		
10.	Connection type	Flange type		
11.	Warranty	Minimum one year warranty period from the date of installation.		

(Note: It is mandatory for the bidders to provide the compliance statement in tabular column format along with catalogue page number (comply/not comply) for the Above points with document proof as required. Failing which bidders will be technically disqualified)