

## CORRIGENDUM - 1

Ref: Tender No. GTB1/LIGY/2021/005/HPLC, Due on 28-06-2021 Up to 3pm

Sl.No	Component	Existing	Amendment
6	Photo Diode Array Detector	4. Slit Width: 1.2 nm and 8 nm. 9. Drift: $5 \times 10^{-4}$ AU/h 10. Flow cell Should be temperature controlled from 5°C above room temperature to 50°C	4. Slit width 1.2 nm and 8 nm or fixed at 6 nm. 9. Drift: $5 \times 10^{-3}$ AU/h 10. This particular point is deleted