

CORRIGENDUM – 1

Tender Specification	Amendment required
Must be able to configurable at least 3 inlet, 4 detectors	Must be able to configurable at least 2 inlet, 3 detectors
Possible to program 15 temp ramps or better	Possible to program 9 temp ramps or better
Must be able to install up to 3 independently temperature controlled injector units simultaneously	Must be able to install up to 2 independently temperature-controlled injector units simultaneously
Automated flow control modules with Pressure set points adjustable in increments of 0.001 psi, settable up to 150psi	Automated flow control modules with Pressure set points adjustable in increments of 0.01 psi, settable up to 100psi
Split ratio up to 7000:1 and suitable for all capillary column from 50um to 530um.	Split ratio up to 6000:1 and suitable for all capillary column from 50um to 530um.
Number of flow programming steps up to 5 or better	Number of flow programming steps up to 3 or better