

	INDIAN INSTITUTE OF TECHNOLOGY MADRAS Chennai 600 036 Telephone: (044) 2257 4436 E-mail: sarathi@ee.iitm.ac.in	
---	---	---

13.11.2020

Department of Electrical Engineering

Corrigendum - 1

Tender Reference no: ELE/SARA/2020/PEF GENERATOR

Name of the Item: High Voltage PEF Generator - 1 Nos

Corrigendum details: Changes in technical specifications

Changes in Technical Specification:

Previous Technical Specifications	Revised Technical Specifications
Output: 100Hz to 15 kHz with rise time of few micro seconds and fall time of about a microsecond	Output: 100Hz to 5 kHz with rise time of few micro seconds and fall time of about a microsecond.
Average output voltage: Maximum variable up to 30kV peak	Average output voltage: Maximum variable up to 30kV peak to peak.

All other terms and conditions remain the same.

Tender Inviting Authority:

Prof. R. Sarathi,
 Department of Electrical Engineering
 IIT Madras
 Chennai - 600 036