

	INDIAN INSTITUTE OF TECHNOLOGY MADRAS Chennai 600 036 Telephone: [044] 2257 9798 E-mail: tender@imail.iitm.ac.in	
---	---	---

24.02.2021

Department of Civil Engineering

Corrigendum-1

Tender Reference no: CE/2021/IOEEOE/040/GASMASS

Name of the Item: Gas Chromatography Mass Spectrometer

Corrigendum details: Changes in technical specification.

Changes in technical specification:

Amendment Ref: CE/2021/IOEEOE/040/GASMASS			
Sl. No	Component	Existing	Amendment
1	General	1. Must be able to configurable at least 3 inlet, 4 detectors.	1. Must be able to configurable at least 2 inlets, 3 detectors.
3	Split/Splitless injector port – 1 No	3. Split ratio: 7000:1 or more.	3. Split ratio: 6000:1 or more.
4	Detector - FID	4. Must have an acquisition time of 2 ms (500Hz) or better.	4. Must have an acquisition time of 2 ms (300Hz) or better.
6	Single Quadruple Mass Spectrometer	3. Electron energy range upto 200eV or better.	3. Electron energy range upto 150eV or better.

Tender Inviting Authority:

The Senior Manager, Project Purchase,
 I.I.T. Madras, Sardar Patel Road, IC & SR Building, 2nd floor,
 Chennai – 600 036