

CORRIGENDUM - 2

Ref: Tender No. GTB1/LIGY/2021/005/HPLC, Due on 28-06-2021 Up to 3pm

Sl.No	Component	Existing	Amendment
6	Photo Diode Array Detector	9. Drift: 5 X 10 ⁻³ AU/h	9. Drift: 1 X 10 ⁻³ AU/h