

CORRIGENDUM – 2

<u>S.No</u>	Tender sub-heading	Specifications	Amendment
1	Injectors	Must be able to install Maximum operating temperature up to 450 °C	Must be able to install Maximum operating temperature up to 400 °C
2	<u>Autosampler</u>	Should able to inject sample volume 1 to 80% of syringe capacity or better Standards should be provided.	Should able to inject sample volume 1 to 50% of syringe capacity or better