

## CORRIGENDUM-1

Ref: Tender No. GTB1/LIGY/2021/001/ATOMIC

Sl. No	Component	Existing	Amendment
6	Optics:	5. Automatic slit selection between 0.2 to 2.0 nm.	5. Automatic slit selection between 0.2 to 1.0 nm
15	Accessories:	1. Types of Cylinders: Acetylene (min 99% pure), Nitrous Oxide (min 99.5% pure) and Argon/Nitrogen with desired purity.	1. Types of Cylinders: Acetylene (1 No; min 99% pure), Nitrous Oxide (1 No; min 99.5% pure) and Argon/Nitrogen (1 No cylinder); with desired purity. <ul style="list-style-type: none"><li>• Suitable purification panels and gas lines should be included.</li><li>• The vendor should take care of the installation of gas lines from the gas cylinders to the instrument.</li><li>• A suitable fume exhaust system should be quoted</li><li>• Must provide single element standards for 24 elements (K, Na, Ag, Al, Au, Bi, Ca, Cd, Cr, Cu, Co, Fe, Mg, Mn, Mo, Ni, Pb, Sb, Si, Te, As, Hg, Se and W)</li><li>• Items to be quoted optionally: Essential consumables and spare parts.</li></ul>