

# **CORRIGENDUM-1**

**Item:** Walk in Environmental Chamber

**Local tender:** CE/MANU/32/IOE23/WALKNCHAMBER

**CORRIGENDUM Details :** Addendum to Technical Specification

## **Addendum to Technical Specification**

### **Item No. 1 Temperature and Humidity Control**

(v): Please assume a total material load inside the chamber of 1000 kg maximum, and minimum temperature ramp rate (up or down) of 5 deg C per hour