

## CORRIGENUDM – 2

**Tender Reference No:** PH/JAYE/02/IOE23/GLOWTOFAB

**Tender Name:** Glovebox Workstation to fabricate LEDs

**Corrigendum details:** Extension of Bid Submission Date and Amendment in Technical Specification

### Extension of Bid Submission Date

The due date for the submission of bids has been extended to 28/03/2023 @ 3 PM. The bid opening is 29/03/2023 @ 3 PM.

### Amendment in Technical Specification

	<b>Existing Specification</b>	<b>Revised Specification</b>
Page No.13&19	Low Temperature Evaporator:	Low Temperature Evaporator: <b>2 in quantity</b>
Page No.13&19	High temperature evaporator:	High temperature Evaporator: <b>2 in quantity</b>
Page No.13&19	Rate Sensor head: Should include 3 nos sensors made from quartz crystal microbalance.(5 or 6 MHz) water-cooled and should include oscillator, feedthrough from base plate.	Rate Sensor head: Should include 3 nos sensors made from quartz crystal microbalance.(5 or 6 MHz) water-cooled and should include oscillator, feedthrough from <b>base plate or top flange</b> .
Page No.14&20	Test boats and crucibles are required to provide during installation	Test boats ( <b>10 in quantity</b> ) and crucibles ( <b>5 in quantity</b> ) are required to provide during installation.