

CORRIGENUDM – 1

Tender Reference No: GTB9/SATYA/2023/14/COMPACTXRD

Tender Name: Chemical Phase analysis by Compact XRD

Corrigendum details: Extension of Bid Submission Date and Amendment in Technical Specification

Extension of Bid Submission Date

The due date for the submission of bids has been extended to 10/03/2023 @ 3 PM. The bid opening is 10/03/2022 @ 5 PM.

Amendment in Technical Specification

Point No	Old Specification	Specification Read as
1.Sample stage		
22	Stage with adjustable z-axis for mount thin pouch cell of 0.5 mm or less for in-operando studies should be provided.	Stage with adjustable z-axis for mount thin pouch cell of 0.5 mm or less.
2.Sample Holders		
27	Sample holder for Coin Cell mounting along with 500 no's of Kapton films should be provided.	Sample holder for Coin Cell mounting along with 500 no's of Kapton films SUITABLE for the coin-cell in-operando studies should be provided.