

CORRIGENUDM – 1

Tender Reference No: GTB10/MANU/IOE23/11/ELECTRODES

Tender Name: pH, Conductivity Meter, Electrodes

Corrigendum details: Amendment in Technical Specification

Amendment in Technical Specification

S.No	Old Specification	Revised Specification
1.	Temperature: -10 to 100 °C	Temperature: 0 to 100 °C