

CORRIGENDUM-2

Item: Walk in Environmental Chamber

Local tender: CE/MANU/32/IOE23/WALKNCHAMBER

CORRIGENDUM Details : Addendum to Technical Specification

Addendum to Technical Specification

Item No. 1 Temperature and Humidity Control

(v): Please assume a total material load inside the chamber of 1000 kg maximum, and minimum temperature ramp rate (up or down) of 5 deg C per hour