

CORRIGENUDM – 2

Tender Reference No: GTB9/SATYA/2023/14/COMPACTXRD

Tender Name: Chemical Phase analysis by Compact XRD

Corrigendum details: Amendment in Technical Specification

Amendment in Technical Specification

Point No	Old Specification	Specification Read as
<u>Goniometer section</u>		
Minimum Step Size		
13	Minimum Step Size: 0.001 deg or better	Minimum step size: 0.005 deg or better