

***Clarification Note for Tender CIE/RAVG/FIST II/007/2013 (18<sup>th</sup> Mar 2013)***

The following clarifications are given for the queries raised after the pre-bid meeting:

- In case a vendor feels that the dynamic performance specification (i.e., stroke, displacement under cyclic loading) for the actuators of the two testing machines cannot be met, the maximum frequency that can be achieved (in lieu of 10 Hz) should be indicated in the technical bid, along with the servovalve details (i.e., type, capacity and number).
- In case a vendor feels that the independent hydraulic service manifold with on/off low and high pressure control and ability to set variable low pressure is not needed for the proper functioning of the machine, an alternative system can be proposed and the corresponding details given in the technical bid.

These clarifications will form part of the tender documents.