## **CORRIGENDUM 1**

## Tender Ref: GTB12/MSR/2023/05/RAMANSPECTRO

Item Name: Micro-photoluminescence and Raman spectrometer

## Reason for Corrigendum: Amendment in Technical Specification

S.NO	Existing	Revised
1	6.C) Microscope stage (Automated including joy stick and computer-controlled)	6.C) Microscope stage (Automated including joy stick and computer-controlled)
	XYZ stage should be provided with travel range for $X = 75$ mm and $Y = 50$ mm or better	XYZ stage should be provided with travel range for $X = 75$ mm and $Y = 50$ mm or better
	XY specifications: repeatability $\leq 1 \ \mu m$ or better; Accuracy $\pm 3 \ \mu m$ or better; Resolution (minimum step size) =10nm or better.	XY specifications: repeatability $\leq 1 \ \mu m$ or better; Accuracy $\pm 3 \ \mu m$ or better; Resolution (minimum step size) = <b>100 nm</b> or better.
	Z specifications: resolution (minimum step size) = $0.01$ micron or better.	Z specifications: resolution (minimum step size) = 0.01 micron or better. "