

Dr. Panchanana Khuntia  
Department of Physics  
Indian Institute of Technology Madras  
Chennai-600036



Tel: +91-44-2257 4847  
Email: [pkhuntia@iitm.ac.in](mailto:pkhuntia@iitm.ac.in)  
Fax: +91-44 2257 4852  
<http://www.physics.iitm.ac.in/~pkhuntia>

Dr. Panchanana Khuntia  
Project Coordinator

Ref.: PHY/2018/010/Stores  
Dated: 06.02.2019

**Limited Tender No.: PHY/2018/010/Stores**

**Due Date: 28.02.2019, 5:00 PM**

## **Corrigendum: Annexure-I**

**Correction: Flange type: KF type**

All other terms and conditions as stated in the original tender dated 06.02.2019 remains the same.

### **SPECIFICATIONS FOR TURBO PUMP WITH ACCESSORIES**

Ultimate Pressure	$<1 \times 10^{-7}$ mbar
Pumping Speed	66 L/s for Nitrogen
Input power	230 VAC and 50 Hz
Flange type	KF type

### **Accessories Required**

The vendor is required to quote the following accessories which are compatible with the above instrument mentioned for purchase together and will be used for price comparison.

Accessories	Quantity
Single Gauge Display Control Unit	1
Venting valve with extension cable	1
Interface Cable	1
Splinter shield	1
Electrical supply cable (>2 m)	1

Active Pirani/Cold Cathode Transmitter	1
Gate valve and manually activated light metal housing	1
Sensor cable (>5 m)	1
Include other essential accessories such as conical adapter, corrugated hose, reducer tee etc. required for the smooth operation of the Turbo Pump	1 each

**N.B.: The successful bidder/vendor should be able to provide excellent life time technical support/service to the turbo pump.**

Sincerely,

Dr. Panchanana Khuntia  
Project Co-ordinator  
Department of Physics  
Indian Institute of Technology Madras  
Chennai-600036, India  
Ph: +91-44-2257 4847  
Email: [pkhuntia@iitm.ac.in](mailto:pkhuntia@iitm.ac.in)