

CORRIGENUDM – 1

Tender Reference No: ME/SAMU/10/IOE23/MATERIALPHASE

Tender Name: Material Phase Characterisation equipment

Corrigendum details: Extension of Bid Submission Date and Amendment to Technical Specification

Extension of Bid Submission Date

The due date for the submission of bids has been extended to 24/04/2023 @ 5 PM. The bid opening is 25/04/2023 @ 3 PM.

Amendment to technical Specification

S.No.	Existing	Revised
1.	Detector: Linear Array CCD 3648 Pixcels or better	Linear Array CCD 2000 Pixels or better
2.	Cooled Detector: yes	Yes (-5C or better, -10C or better)
3.	Slit Size: Continuous 10µm to 400 µm variable	25 micrometer or better
4.	Spectrometer FL: 200 mm or better	100 mm or better
5.	Exposure time: 1ms to 60 sec or better	10 ms to 60 sec or better
6.	Laser: Wavelength: 532nm/785nm Power: 300mW or higher Stability: 1 % or better	Laser: Wavelength: 532nm/785nm Power: 100mW/300mW or higher Stability: 1 % or better