

CORRIGENDUM

Date:20th May, 2019

Subject: Limited tender for thermal evaporation system

Reference: ELE/DEBD/2019/THERMAL EVAPORATION SYSTEM

This corrigendum is in reference to the tender released on 1st May, 2019 for thermal evaporation system. Following are the changes made to the current tender.

1. The tender closing date is extended to **31st May, 2019**.
2. Under Section 1, Non-magnetic and powder coated chamber is not necessary. Chamber can have maximum dimensions of **350mm (Ø) X 420mm (H)**.
3. Under section 4(a), specially designed temperature controllable sources should have temperature controllability of **±1°C**
4. Under section 4(a), two power supplies (200A) are required for the LT electrolytic copper wire sources and two temperature controllers are required for the specially designed temperature controllable (temperature controllability of **±1°C**) sources.