

Dated:13.11.2018

CORRIGENDUM FOR THE SPECIFICATION OF FABRICATION MACHINE

Reference:

CHY/KOTH/2019-20/006/SPLX

Corrigendum-I

1. Supplier should provide atleast five proof of installation of the similar or higher-end glove box (same make as what vendor intent to supply to us) in India along with the user details. Most importantly the glove box should be suitable for Li-ion battery research maintaining very low moisture and oxygen content (around 0.1ppm for most part of the time, but not exceeding 1ppm) (OR) End-user certificate mentioning the purchase date, the current status of the instrument and its suitability to Li-ion battery research.
2. For other items, proof of installation in India at least in 3 places is required.
3. A detailed compliance certificate should be included in the technical bid along with the relevant web and catalog materials.



Signature

Prof. Kothandaraman Ramanujam