



INDIAN INSTITUTE OF TECHNOLOGY MADRAS
Chennai 600 036

Telephone: (044) 2257 4230

E-mail: sghosh@iitm.ac.in



Prof. Sundargopal Ghosh
Project Coordinator

Ref: CHY/SGHO/2019-20/0015/SPLX

Dated: 20.03.2020

Limited Tender No: CHY/SGHO/2019-20/0015/SPLX

Due Date: 09.04.2020, 3:00pm

Pre-Bid meeting: - Not required.

Bid opening meeting on Due Date: 09.04.2020, 4:00pm

Dear Sir/Madam,

On behalf of the Indian Institute of Technology Madras, offers are invited for the supply of “**LOW TEMPERATURE CIRCULATORY CHILLER**” conforming to the specifications given in (**Annexure-I**).

Terms and Conditions of Limited Tender

- 1. Preparation of Bids:** - The Limited tenders should be submitted under **one bid system** (i.e.) Technical-cum-Financial bid.
- 2. Delivery of the tender:** - The tender shall be sent to the below-mentioned addresses either by post or by courier (duly sealed and super scribed on the envelope with the reference No and due date) so as to reach the following address before the due date and time specified in our Schedule:
Prof. Sundargopal Ghosh,
Department of Chemistry,
IIT Madras
Chennai - 600 036
- 3. Price:** - The price should be quoted in net per unit (after breakup) and must include all packing and delivery charges to **Department of Chemistry**.
 - a. The offer/bid should be exclusive of taxes and duties. The percentage of tax & duties should be clearly indicated separately. IIT Madras is eligible for concessional GST and relevant certificate will be issued.
 - b. In case of import supply, the price should be quoted without custom duty. IIT Madras is exempted from levy of IGST on Imports and eligible for concessional custom duty (not exceeding 5%) and the price should be quoted on EX-WORKS and CIP (stating the Cost, Insurance, Freight separately) and indicating the mode of shipment.
- 4. Terms of Delivery:** - The item should be supplied to our Departments as per Purchase Order. In case of import supply, the item should be delivered at the cost of the supplier to our Institution. The Installation/Commissioning should be completed as specified in our important conditions.
- 5. Catalogue:** Original catalogue (not any photocopy) of the quoted model duly signed must accompany the quotation in the Technical-cum-financial bid
- 6. Late offer:** - The offers received after the due date and time will not be considered

7. **Payment:** - No Advance payment will be made for Indigenous purchase. However, 90% Payment against Delivery and 10% after installation are agreed to wherever the installation is involved. In case of import supplies the payment will be made only through 100% Letter of Credit i.e. (90% payment will be released against shipping documents and 10% after successful installation wherever the installation is being done).
8. **Advance Payment:** - No advance payment is generally admissible. In case of specific percentage of advance payment is required, the Vendor has to submit a Bank Guarantee from a Nationalized Bank of India equal to the amount of advance payment.
9. **On-site Installation:** - The equipment or machinery has to be installed or commissioned by the successful bidder within number of days (as prescribed by PI's) from the date of receipt of the item at site of IIT Madras.
10. **Warranty/Guarantee:** - The offer should clearly specify the warranty or guarantee period for the machinery/equipment.
11. **Validity:** Validity of Quotation not less than 60 days from the due date of tender
12. **Bid Opening:** The bid will be opened on **09.04.2020, 4:00pm at the Department of Chemistry, IIT Madras.**
13. **Accept /Reject:** IIT Madras reserves the full right to accept / reject any tender at stage without assigning any reason.
14. **Settlement of Disputes:** Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Chennai in Tamil Nadu.
15. **Risk Purchase Clause:** - In the event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from other sources on the total risk of the supplier under risk purchase clause.
16. **Unsolicited offers:** "This notice is being published **for information only** and is not an open invitation to quote in this limited tender. Participation in this tender is by invitation only and is limited to the selected registered suppliers. Unsolicited offers are liable to be ignored. However, suppliers who desire to participate in such tenders in future may apply for registration as per procedure." The Website for Registration of vendors is <http://web.iitm.ac.in/supplier/> and the mail address for queries is "workflow@rt.iitm.ac.in".

Yours sincerely,

**Prof. Sundargopal Ghosh,
Department of Chemistry,
IIT Madras
Chennai - 600 036**

SPECIFICATIONS FOR LOW TEMPERATURE CIRCULATORY CHILLER

- a. Temperature range : - 85 to + 100 deg C
- b. Temperature Stability : +/- 0.05 Deg C
- c. Temperature Display : TFT LCD
- d. Display resolution : 0.1
- e. Temperature Selection : Touch sensitive Resistive
- f. Temperature calibration : minus 60, 0, and +0 Deg C
- g. Alarm Indication : High/ Low temperature
- h. Temperature Sensor : Pt-100 DIN 1/3 Class A
- h. Auto start : Provided. Enable parameters call back
- I. Refrigerant CFC Free : EPA (USA) Approved Dupont and R23 Refrigerant
- J. True PID : yes, the R23 Compressor is not switched OFF and On
- k. Control Logic : PWM driven Valve Control in By-Pass mode.
- l. Operating Panel : TFT , Splash proof
- m. Cooling capacity : 1200 W at + 20 deg C
900 W at – 20 deg C
500 W at – 40 deg C
200 W at – 80 deg C
- o. External Pt 100 sensor : 150 mm x 6 mm ID, cable 2 m
- p. Dimension L x W x H cm : 45 x 31 x 60 cm
- q. Weight : 95 kg
- r. Power supply : 3 Ph 400 V /50 Hz
- s. Wheels for easy movement : castor wheels to be provided
- t. Delivery : Free of Cost at IIT Chennai
- u. Installation/demonstration : Free of cost by SISKIN Engineers
- v. Warranty : 24 months against any manufacturing defect
- w. Certification : CE & ISO
- x . Manufacturer : NSIC/MSME Registered