

Minutes of Pre-bid Meeting

Tender No : OED/2016/003 to 011/Project/VSUN
For procurement of : for procurement of various lab equipments
Venue : Seminar Hall, Ocean Engg., IIT, Chennai.
Date & Time : 03/02/2017 & 4.00p.m


The following participants attended the meeting:

Representatives of IITM:

- 1. Dr.S.A.Sannasiraj** - H.O.D
- 2. Dr.V.Sundar** - Project Co-Ordinator
- 3.T.Sarvaran** - Asst.Registrar

Representatives of Prospective Bidder's:

Sl.No.	Name of Firm	Represented by
1.	Indus Instrument	C.Bala Krishnan
2.	Elcome Technologies Pvt. Ltd	Sanal Jagadishan
3.	Elcome Integrated System Pvt. Ltd	T.Janagan
4.	Pan India Consultants Pvt Ltd	C.Mohan
5.	AIMIL Ltd.,	K.Kannan



Minutes of Pre-bid Meeting

Sl.No.	Query from	Gist of Query	Response
1.	AIMIL Ltd.,	Instrument: Total Station with GPS Clarify the point- Is it necessary to have in-built camera since it has no usage in data collection purpose.	Items may be quoted with in-built camera and without in-built camera separately.
2.	AIMIL Ltd.,	Instrument: Total Station with GPS Clarify the point- Whether Rechargeable batteries can be supplied in 2 nos. which have 6hr + 6hr battery backup?	The provision of two batteries should be made such that the measurements are not interferred.
3.	Indus Instruments	Instrument: DGPS Clarify the point- Is it possible to have external Bluetooth connectivity system, apart from integrated Bluetooth system.	Acceptable provided satisfying the required specifications.
4.	a)Elcome Technologies Pvt.Ltd., b)Pan India Consultants Pvt. Ltd.	Instrument: DGPS Clarify the Point- No. of Channels : >200channels in technical specifications and whether it is mandatory or not.	Radio Frequency is not Mandatory, if license is required
5.	a)Elcom Technologies Pvt.Ltd., b)Elcome Integrated Systems	Instrument: Total Station and DGPS. Clarify the Point- OGC Compliance certificate needed.	It is Mandatory to confirm the Design Standard of instrument. (Ref: Open Geospatial Consortium))

N. Sridhar