

## Minutes of Pre-bid Meeting

**Tender No** : OED/2017/011/PROJECT/VSUN

**For procurement of** : For Procurement of MOTION SENSOR (Heave Sensor)

**Venue** : Room No-305, 3rd Floor, Department of Ocean Engineering, IIT Madras.

**Date & Time** : 7<sup>th</sup> February 2018 @ 3.00 pm

The following participants attended the meeting:

### **Representatives of IIT:**

- |                         |              |
|-------------------------|--------------|
| 1. Prof.S.SURENDRAN     | Chairman     |
| 2. Prof.V.SUNDAR        | Co-ordinator |
| 3. Prof.S.A.SANNASIRAJ  | Member       |
| 4. Prof.K.MURALI        | Member       |
| 5. Prof.ANANTHAKRISHNAN | Member       |

### **Representatives of Prospective Bidder's:**

- | Sl.No. | Name of Firm  | Represented by                    |
|--------|---------------|-----------------------------------|
| 1.     | Mr.T. Janagan | Elcome Integrated System Pvt. Ltd |

*V. Sundar*



## Lists of Corrigendum

SL.No.	For Limited Tender	Existing Description	To be read as
1.	OED/2017/011/ PROJECT/VSUN	<b>4.Electrical:</b> Voltage Input 09 to 36 V DC	<b>Electrical:</b> Voltage Input - <b>Between</b> 09 to 36 V DC
2.		<b>5.Serial Ports:</b> Serial RS 232 / 422 -Serial input / 3 serial out put Ethernet - 5 serial port	<b>Serial Ports:</b> Serial RS 232 / 422 <b>&amp; Ethernet-</b> 5 Serial input / 3 serial out put
3.		<b>5.Serial Ports:</b> Output Rate - 200 Hz	<b>Serial Ports:</b> Output Rate – <b>1 to 200 Hz</b>
4.		<b>5.Serial Ports:</b> Logging Capacity - 8 GB of 48H	<b>Serial Ports:</b> <b>Interface</b> Logging Capacity - 8 GB of 48H
5.		<b>6.Environmental Specifications:</b> Depth Rating - Upto 200 m	<b>Environmental Specifications:</b> Depth Rating - Upto 200 m <b>(Bathymetry survey Operation)</b>

*N. Sundar*