

Tender submission due date extended up to 3.00 pm on 25.02.2019

Tender No: CIE/Ligy/001/2019

Name of the Item: High Performance Liquid Chromatography

Minutes of the pre-Bid meeting held on 01-02-2019 for the above subject is appended below

Specifications

Sl. No.	Items	Section	Points raised by the vendor	IITM clarifications
1	HPLC	General features of HPLC	(i)...“manual injector” to be removed (ii)...“optional auto sampler..” to be changed to “auto sampler”	Committee accepted both the suggestions
2	HPLC	Pump	Flow rate setting range as “0.01 to 5 mL” instead of 0.01 to 10 mL	Committee decided to retain the original specification
			Maximum pressure range 5000 psi instead of 8000 psi	Committee decided to retain the original specification
			Delay volume has to be mentioned	Committee accepted the suggestion. Delay volume: 1 mL or less
3	HPLC	UV-Vis detector	Wavelength range “190-600 nm or better” instead of “190-800 nm or better”	The committee decided to modify the original specification as “Wavelength Range: 190 - 700 nm or better”.
4	HPLC	Columns and consumables	HPLC columns specification to be mentioned	Committee decided to include the following HPLC analytical column specification. C18: 250 mm × 4.6 mm × 5 μ C18: 250mm × 4.6 mm × 3 μ
5	HPLC	Auto sampler: Injection volume range	“...with 200 μL loop..” to be removed	Committee accepted the suggestion

Minutes of Pre-bid meeting held on 1-2-2019 for above is appended below:

TENDER NOTICE

Item No.̀	Section No	Points raised by Vendor	IITM clarifications
Tender Notice Instructions to the Bidder	Schedule Payment	Vendors requested for 100% LC advance payment against Performance guarantee	ICSR decided to retain the original Terms and Conditions

GENERAL TERMS AND CONDITIONS

Item No.̀	Section No	Points raised by Vendor	IITM clarifications
General Terms and Conditions	Technical Bid should comprise of the following:	EMD should not be enclosed in the Technical Bid	The EMD should be sent in a sealed cover