

## INDIAN INSTITUTE OF TECHNOLOGY MADRAS Chennai 600 036

TUV NORD
TH India Product LEL

SO 9001

Telephone: [044] 2257 9763 E-mail: tender@imail.iitm.ac.in

The Manager (Project Purchase)

Date: 15.02.2023

Open Tender Reference No: AE/SUJI/06/IOE23/DMFCON

GEM NAR ID: GEM/GARPTS/13022023/CD8HY0JIUSVZ Due Date/Time: 01.03.2023@ 3:00 PM

Dear Sir/Madam,

On behalf of the Indian Institute of Technology Madras, Tenders are invited in two bid system from Class-I local suppliers and Class II local suppliers, for the supply of "Digital Mass Flow Controller" Conforming to the specifications given in Annexure -A.

Tender Documents may be downloaded from Central Public Procurement Portal <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a>. Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a>. The portal enrolment is free of cost. Bidders are advised to go through instructions provided at "Help for contractors". [Special Instructions to the Contractors/Bidders for the e-submission of the bids online through this eProcurement Portal"]

Bidders can access tender documents on the website (For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, click on "GO" button to view all IIT Madras tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a> as per the schedule attached.

NA
Vendor registration code. Vendor registration with IC&SR (IITM) is mandatory for bidders to participate in tenders.  ** For Vendor Registration & Guidelines, Please follow the website: https://icandsr.iitm.ac.in/vendorportal; Helpdesk: vendorhelpdesk@icsrpis.iitm.ac.in

**No manual bids will be accepted.** All tender documents including Technical and Financial bids should be submitted in the E-procurement portal.

Last date for receipt of tender	:	01.03.2023 @ 3:00 PM
Date & time of opening of tender	:	02.03.2023 @ 3:00 PM

## 3. Instructions to the Bidder:

<u>A)</u>	Searching for tender documents	<ul> <li>There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, organization name, location, date, value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.</li> <li>Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective "My Tender" folder. This would enable the CPP Portal to intimate the bidders through SMS / email in case there is any corrigendum issued to the tender document.</li> <li>The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.</li> </ul>
<u>B)</u>	Assistance to bidders:	<ul> <li>Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.</li> <li>Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is [0120-4200462, 0120-4001002, 0120-4001005]</li> </ul>
<u>C)</u>	Enrollment Process to: Bidders	<ul> <li>Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal <u>URL:https://etenders.gov.in/eprocure/app</u> by clicking on "Online Bidder Enrollment". Enrollment on the CPP Portal is free of charge.</li> <li>As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.</li> <li>Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.</li> <li>Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.)</li> <li>Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse.</li> <li>Bidder then may log in to the site through the secured log-in by entering their user ID / password and the password of the DSC / eToken.</li> <li>Possession of a Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/e-token in the company's name is a prerequisite for registration and participating in the bid submission activities through https://etenders.gov.in/eprocure/app</li> </ul>

## Digital Signature Certificates can be obtained from the authorized certifying agencies, details of which are available in the web site https://etenders.gov.in/eprocure/app under the "Information about DSC". Preparation of bids D) Bidder should take into account any corrigendum published on the tender document before submitting their bids. Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Please note the number of covers in which the bid documents have to be submitted, the number of documents including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid. Bidder, in advance, should prepare the bid documents to be submitted as indicated in the tender document / schedule and generally shall be in PDF XLS formats as the case may be. Bid documents may be scanned with 100 dpi with black and white option. To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, GSTIN Details, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My Documents" area available to them to upload such documents. These documents may be directly submitted from the "My **Documents**" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process. Submission of bids E) Bidder should log into the site well in advance for bid submission so that he/she can upload the bid in time i.e. on or before the bid submission date and time. Bidder will be responsible for any delay due to other issues. The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document. Bidder has to select the bid security declaration. Otherwise, the tender will be summarily rejected. A standard BOQ format has been provided with the tender document to be filled by all the bidders. Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. Bidders are required to download the BOQ file, open it and complete the detail with their respective financial quotes and other details (such as name of the bidder). If the BOQ file is found to be modified by the bidder, the bid will be rejected. The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission. The Tender Inviting Authority (TIA) will not be held responsible for any sort of delay or the difficulties faced during the submission of bids online by the bidders due to local issues. The uploaded tender documents become readable only after the tender opening by the authorized bid openers.

		<ul> <li>Upon the successful and timely submission of bids, the portal will give a successful bid submission message &amp; a bid summary will be displayed with the bid no. and the date &amp; time of submission of the bid with all other relevant details.</li> <li>Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.</li> <li>More information useful for submitting online bids on the CPP Portal may be obtained at: <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a>.</li> <li>All tender documents including pre-qualification bid, Technical Bid &amp;Financial Bid should be submitted separately in online CPP portal as per the specified format only. Right is reserved to ignore any tender which fails to comply with the above instructions. No manual bid submission will be entertained.</li> </ul>
<u>F)</u>	Marking on Technical Bid	• The bidder eligibility criteria, technical specification and supply of item for this tender is given in Annexure A.
		• The Bidders shall go through the specification and submit the technical bid.
		• The Technical bid should be submitted in the proforma as per Annexure-B in pdf format only through online (e-tender). No manual submission of bid will be entertained.
		• The technical bid should have a page-wise heading as "Technical Bid" and page no. in all pages with seal and signature of authorized signatory. The total no. of pages should be mentioned at the last page of the documents.
		• The technical bid should consist of bidder eligibility criteria details and all technical details along with catalogue/ pamphlet which will give a detailed description of product with technical data sheet so that technical compliance can be verified.
<u>G)</u>	Marking on Price Bid	• Financial bid (BoQ) should be submitted in the prescribed proforma format as per Annexure-C in xls format through e-tender only. No manual or other form of submission of Financial Bid will not be entertained

4) **Preparation of Tender**: The bidders should submit the bids in two bid system as detailed below.

## Bid I \_Technical Bid

The technical bid should consist of bidder eligibility criteria and technical specification compliance sheet as per Annexure-B.

### Bid II \_Price Bid

The price bid should be submitted in excel format (BoQ) as per the proforma (Annexure C) uploaded in the e-Tender web site. The Quoted price should be for supply and installation of the item and inclusive of all cost and statutory levies at IIT Madras.

## 5) Price:

- a) The price should be quoted only in INR net per unit (after breakup) and must include all packing, transit insurance and delivery charges to the **Department of Aerospace Engineering**.
- b) The rate quoted shall be all inclusive of all taxes and no extra payment will be made other than statutory revisions as per the terms and conditions stipulated in this contract document.
- c) The percentage of tax & duties should be clearly indicated separately. IIT Madras is eligible for custom duty (5.5%). Relevant certificates will be issued wherever necessary.

d) The offer/bids should be submitted through online only in two bid system i.e. Technical Bid and Financial Bid separately.

## 6) Tenderer shall submit along with this tender:

- (i) Proof of having ISO or other equivalent certification given by appropriate authorities.
- (ii) Name and full address of the Banker and their swift code and PAN No. and GSTIN number.
- (iii) GST registration proof showing registration number, area of registration etc.
- (iv) All of your future correspondences including Invoices should bear the GST No. and Area Code.

### 7) Terms of Delivery:

Supplier will be fully responsible for the safe carriage, Installation/Commissioning of goods up to the **Department of Aerospace Engineering**., IIT Madras or named place as per PO, Insurance coverage will be in the scope of the supplier.

The tenderer should indicate clearly the time required for delivery of the item (subject to the approval of the Executive Committee-IIT-Madras). In case there is any deviation in the delivery schedule, liquidated damages clause will be enforced or penalty for the delayed supply period will be levied.

In the event of delay or non-supply of materials/execution of Contract beyond the date of delivery/completion of job. The penalty will be levied @1% per week of delay subject to a max of 10% of the value of purchase order and if the delay is more than accepted time frame by IIT M, the PO would be partially or fully cancelled and liquidated damages will be enforced accordingly.

## 8) Period for which the offer will remain open:

The Tender shall remain open for acceptance/validity till: 120 days from the date of opening of the tender. However, the day up to which the offer is to remain open being declared closed holiday for the Indian Institute of Technology Madras, the offer shall remain open for acceptance till the next working day.

#### 9) **EMD**:

The EMD of **Rs.60,000** to be transferred to the account details mentioned in Annexure D and proof should be enclosed in the Technical Bid. Any offer not accompanied with the EMD shall be rejected summarily as non-responsive.

The EMD of the unsuccessful bidders shall be returned within 30 days of the end of the bid validity period. The same shall be forfeited, if the tenderers withdraw their offer after the opening during the bid validity period. The Institute shall not be liable for payment of any interest on EMD.

EMD is exempted for Micro and Small Enterprises (MSE) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium Enterprises (MSME) and Startups as recognized by Department of Industrial Policy & Promotion (DIPP). (MSE/MSME/DIPP PROOF should be enclosed in the cover containing technical bid).

### 10) Performance Security: -

The successful bidder should submit Performance Security for an amount of 3% of the value of the contract/supply. The Performance Security may be furnished in the form of an Account Payee DD, FD Receipt in the name of "The Registrar, IIT Madras" from any scheduled commercial bank or Bank Guarantee from any scheduled commercial bank in India. The performance security should be furnished within 14 days from the date of the purchase order.

Performance Security in the form of Bank Guarantee: - In case the successful bidder wishes to submit Performance Security in the form of Bank Guarantee, the Bank Guarantee should be routed directly to IIT Madras from the Bank.

The Bank Guarantee should remain valid for a period of sixty days beyond the date of completion of all contractual obligations of the supplier including the warranty obligations.

- For the same tender, either the OEM or the authorized dealer/service provider can only quote. But both of them cannot quote separately for the same tender.
- The offers/bids should be sent only for a item/Equipments of latest version that is available in the market and supplied to a number of customers. A list of customers in India with details must accompany the quotations. Quotations for a prototype machine will not be accepted
- Original catalogue (not any photocopy) of the quoted model duly signed by the principals must accompany the quotation in the Technical bid.
- Compliance or Confirmation report with reference to the specifications and other terms & conditions should also be obtained from the principal/OEM.

#### 15) Risk Purchase Clause

In the event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from other sources on the total risk of the supplier under risk purchase clause.

## 16) Payment:

- (i) No Advance payment will be made. However, 90% Payment against Delivery and 10% after installation are agreed to wherever the installation is involved.
- (ii) Advance Payment: No advance payment is generally admissible. In case a specific percentage of advance payment is required, the Vendor has to submit a Bank Guarantee from a scheduled commercial bank in India equivalent to the amount of advance payment.

## 17) On-site Installation:

The equipment/item or Machinery has to be installed or commissioned by the successful bidder within the number of days (as prescribed by PI) from the date of receipt of the item at the site of IIT Madras.

## 18) Warranty/Guarantee:

The offer should clearly specify the warranty or guarantee period for the machinery/equipment. Any extended warranty offered for the same has to be mentioned separately (For more details please refer our Technical Specifications).

\*\* Note: PO which involves installation, warranty/guarantee shall be applicable from date of installation.

## 19) Acceptance and Rejection:

Failure to comply with any of the instructions stated in this document or offering unsatisfactory explanations for non-compliance will likely to lead to rejection of offers.

I.I.T. Madras has the right to accept the whole or any part of the Tender or portion of the quantity offered or reject it in full without assigning any reason.

## 20) Debarment from Bidding:

In case of breach of Terms & Conditions, Bidder may be suspended from being eligible for bidding in any contract with the IIT Madras up to 2 Years [as per Rule 151(iii) of GFR] from the date of Tender.

## 21) Disputes and Jurisdiction:

**Settlement of Disputes:** Any dispute, controversy or claim arising out of or in connection with this PO including any question regarding its existence, validity, breach or termination, shall in the first instance be attempted to be resolved amicably by both the Parties. If attempts for such amicable resolution fails or no decision is reached within 30 days whichever is earlier, then such disputes shall be settled by arbitration in accordance with the Arbitration and Conciliation Act, 1996. Unless the Parties agree on a sole arbitrator, within 30 days from the receipt of a written request by one Party from the other Party to so agree, the arbitral panel shall comprise of three arbitrators. In that event, the supplier will nominate one arbitrator and

the Project Coordinator of IITM shall nominate on arbitrator. The Dean IC&SR will nominate the Presiding Arbitrator of the arbitral tribunal. The arbitration proceeding shall be carried out in English language. The cost of arbitration and fees of the arbitrator(s) shall be shared equally by the Parties. The seat of arbitration shall be at IC&SR IIT Madras, Chennai.

- a. **The Applicable Law:** The Purchase Order shall be construed, interpreted and governed by the Laws of India. Court at Chennai shall have exclusive jurisdiction subject to the arbitration clause.
- b. Any legal disputes arising out of any breach of contact pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Chennai in Tamil Nadu.
- **Force Majeure:** The Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.

For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.

If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.

## Eligibility Criteria:

- As per the Government of India Order, only "Class I Local Suppliers" and "Class II Local Suppliers" can participate in this tender.
- Bidder should confirm their acceptance that they comply with the provisions with report to "Guidelines for eligibility of a bidder from a country which shares a land border with India as detailed at Annexure-F. The bidder should submit Certificate for "Bidder from/ Not from Country sharing Land border with India & Registration of Bidder with Competent Authority" as per Order of DoE F.No.6/18/2019-PPD dated 23.07.2020 as mentioned.
- Preference to "class I Local Suppliers": preference will be given to "class 1 local suppliers" (subject to class -I local supplier's quoted price falling within the margin of purchase preference ) as per public procurement (preference to make in India) order 2017 .O.M No P- 45021/2/2017 pp(BE 11) dt 04/06/2020 subject to the conditions that the "class 1 Local Supplier" should agree to supply goods / provide service at L1 rate and furnish a certificate with the technical bid document that the goods/service provided by them consists local content equal to or more than 50%.( certificate from Chartered Accountant in case value of contract exceeds Rs 10 crore).
  - <sup>2</sup> 'Class I local supplier' means a supplier or service provider whose goods, services or works offered for procurement consists of local content equal to or more than 50% as defined under the above said order. Declaration to be provided as per Annexure-E per item/service/work.
  - Class II local supplier' means a supplier or service provider whose goods, services or works offered for procurement consists of local content equal to 20% but less than 50% as defined under the above said order. Declaration to be provided as per Annexure-E per item/service/work.
  - Margin of purchase preference': The margin of purchase preference shall be 20%. The Definition of the margin of purchase preference is defined in the Govt. of India Order No: P-45021/12/2017-PP (BE-II) Dt.4th June, 2020) Order 2017. As per the Government of India Order "Margin of Purchase Preference" means the maximum extent to which the price quoted by a "Class-I local supplier" may be above the L1 for the purpose of purchase preference.

\*\*Note: Local content percentage to be calculated in accordance with the definition provided at clause 2 of revised public procurement preference to Make in India Policy vide GoI Order no. P-45021/2/2017-PP (B.E.-II) dated 15.06.2017 (subsequently revised vide orders dated 28.05.2018,

	29.05.2019and 04.06.2020) MOCI order No. 45021/2/2017-PP (BE II) Dt.16th September 2020 & P-45021/102/2019-BE-II-Part(1) (E-50310) Dt.4th March 2021
25)	Evaluation of Bids
<b>2</b> 3)	Bid evaluation will take place in two stages.
	Stage I Technical Bid evaluation
	All bidders who have fully complied with bidder eligibility criteria I, II and technical evaluation
	(Annexure A) will only be considered for opening of price bid.
	Stage II: Price Bid Evaluation
	The price bid evaluation will be based on price quoted by the bidder. The rate quoted for "Digital Mass
	Flow Controller" quantity mentioned in Annexure-A will alone be taken up for arrival of Lowest Bid
	(L1) value.
26)	Selection of successful bidder and Award of Order
_0)	The order will be directly awarded to the technically qualified bidder as per the condition in para 3A of
	DIPP, MoCI Order No. 45021/2/2017-PP (BE II) dated 16th September 2020.
27)	All information including selection and rejection of technical or financial bids of the prospective bidders
21)	will be communicated through e-Tender portal. In terms of Rule 173(iv) of General Financial Rule 2017,
	the bidder shall be at liberty to question the bidding conditions, bidding process and/or rejection of bids.
28)	The tenderer shall certify that the tender document submitted by him / her are of the same replica of the
<b>40</b> )	tender document as published by IIT Madras and no corrections, additions and alterations made to the
	same. If any deviation found in the same at any stage and date, the bid / contract will be rejected /
	terminated and actions will be initiated as per the terms and conditions of the contract.
20)	Due to Covid-19 pandemic pre-bid meeting will be conducted through online. Clarification to the queries
<b>29</b> )	and doubts raised by the bidders will be issued as a corrigendum/addendum in the e-tenders portal.
20)	Due to Covid-19 pandemic the bidders will not be entertained to participate in opening of Bids. Since the
30)	tender is e-tender, the opening of the bids may be checked using the respective logins of the bidders.

## **ACKNOWLEDGEMENT**

It is hereby acknowledged that I/We have gone through all the points listed under "Specification, Guidelines, Terms and Conditions" of tender document. I/We totally understand the terms and conditions and agree to abide by the same.

SIGNATURE OF TENDERER ALONG WITH SEAL OF THE COMPANY WITH DATE

## Bidder Eligibility Criteria and Technical Specification for "Digital Mass Flow Controller" Tender No. AE/SUJI/06/IOE23/DMFCON

### Bidder Eligibility Criteria – I (Public Procurement – Preference to Make in India)

Only 'Class-I local suppliers' and 'Class-II local suppliers', as defined under DIPP, MoCI Order No. P-45021/2/2017-PP (BE-II) dated 16<sup>th</sup> September 2020 and other subsequent orders issued therein.

## Bidder Eligibility Criteria – II

The bidder/OEM should have supplied at least 3 similar items to IITs, NITs, IISERs, CSIR Labs or other Govt. R&D organizations in the last 5 years, PO copies or installation certificates along with contact details of end user need to be submitted as the proof of supply. IIT Madras reserves its right to verify the claims submitted by the bidder and the feedback from the previous customers will be part of technical evaluation.

## III. Technical Specification for "Digital Mass Flow Controller"

S.No	Equipment name	Requirement	Quantity
1		Flow Range: 2000 SLPM	1
2		Flow Range: 1000 SLPM	2
3	Distract Advantage (AAFC)	Flow Range: 500 SLPM	2
4	Digital Mass Flow Controller (MFC)	Flow Range: 250 SLPM	1
5		Flow Range: 100 SLPM	2
6		Flow Range: 50 SLPM	1

S.No	Technical Specification	
1	All the Mass Flow controllers should be of multi-gas selectable type. The Non-Corrosive Mass flow controllers should be calibrated for gases: H2, LPG, N2, Ar, O2, He, Air, CH4, CO, CO2, C2H2	
2	MFC should have the option for Personalized gas mixing compositions for accurate mixed gas measurement and switch between selected gases on the fly without any need for recalibration of the mass flow controller.	
3	Operating Pressure: Max 160 PSIA	
4	Temperature range of operation: -10 ° C to 50 ° C	
5	Wetted and Seal Material: Recommended Body material and seals suitable for the gas (preferably SS302)	
6	Accuracy: ± 1% of actual value or (0.8% of Reading ± 0.2% of Full Scale)	
7	Pressure sensitivity: ± 0.01% of full scale / PSI	
8	It should have rigid metallic construction	

9	The output of the controller should be 0 – 5 V or 4 - 20 mA signals				
10	The controller should be provided with circuit protection				
11	It should be operated in the po	wer supply of 230 V AC, 50 Hz			
12	The controller should have the	connection of compression gas fittings			
13	Setting Flow Ramp up program keypad with the display on the	ming cycle should be possible by interfacing through the in build MFC			
14	Typical Response Time <100 ms	s (Adjustable)			
15	Control Range : 0.01% to 100%	Full Scale			
16	Electrical Connection Options:	DB9 Pin or RS 232			
17	Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices				
18	All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter ) should be supplied along with the instrument				
19	Detailed service manual and operating manual with circuit diagram should be provided along with the instrument				
20	Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.				
21	Manufacturers should have their own Re-Calibration and Service Centre in India.				
ADDITION	ADDITIONAL TERMS AND CONDITIONS				
22	Warranty	1 year from the date of installation			
23	Installation and service	Installation, Commissioning, Training and Service should be done by the supplier			

Note: All the 6 Digital Mass Flow Controller(MFC) Should be from the same OEM

## TECHNICAL BID PROFORMA Tender No. AE/SUJI/06/IOE23/DMFCON

**Item Name: "Digital Mass Flow Controller"** 

**Bidder Eligibility Criteria:** 

1.0	Bidder Eligibility Criteria-I (Public Procurement – Preference to Make in India)	Class I / Class II	Local Content value	Reference, Page No.
I	Only 'Class-I local suppliers' and 'Class-II local suppliers', as defined under DIPP, MoCI Order No. P-45021/2/2017-PP (BE II) dated 16 <sup>th</sup> September 2020 and other subsequent orders issued therein.			
2.0	Bidder Eligibility Criteria-II	Compliance (Yes/No)	Reference Page No.	Remarks, If any
1	The bidder/OEM should have supplied at least 3 similar items to IITs, NITs, IISERs, CSIR Labs or other Govt. R&D organizations in the last 5 years, PO copies or installation certificates along with contact details of end user need to be submitted as the proof of supply. IIT Madras reserves its right to verify the claims submitted by the bidder and the feedback from the previous customers will be part of technical evaluation.			

## 3.0 **Technical Compliance:**

S.No	Equipment name	Requirement		Complied / Not Complied	Reference Page No.
1		Flow Range: 2000 SLPM	1		
2		Flow Range: 1000 SLPM	2		
3	Digital Mass Flow Controller (MFC)	Flow Range: 500 SLPM	2		
4		Flow Range: 250 SLPM	1		
5		Flow Range: 100 SLPM	2		
6		Flow Range: 50 SLPM	1		

S.No	Technical Specification	Complied / Not Complied	Reference Page No.
1	All the Mass Flow controllers should be of multi-gas selectable type. The Non-Corrosive Mass flow controllers should be calibrated for gases: H2, LPG, N2, Ar, O2, He, Air, CH4, CO, CO2, C2H2		

MFC should have the option for Personalized gas mixing compositions for accurate mixed gas measurement and switch between selected gases on the fly without any need for recalibration of the mass flow controller.  3			
4 temperature range of operation: -10 °C to 50 °C  5 Wetted and Seal Material: Recommended Body material and seals suitable for the gas (preferably S3302)  6 Accuracy: ± 1% of actual value or (0.8% of Reading ± 0.2% of Full Scale)  7 Pressure sensitivity: ± 0.01% of full scale / PSI  8 It should have rigid metallic construction  9 The output of the controller should be 0 – 5 V or 4 - 20 mA signals  10 The controller should be provided with circuit protection  11 It should be operated in the power supply of 230 V AC, 50 Hz  12 The controller should have the connection of compression gas fittings  13 Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14 Typical Response Time <100 ms (Adjustable)  15 Control Range: 0.01% to 100% Full Scale  16 Electrical Connection Options: D89 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17 Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided do now their own Re-Calibration and Service  Necessary accessories such as Power Supply, Communication cable, softwares should have their own Re-Calibration and Service	2	for accurate mixed gas measurement and switch between selected gases on the fly without any need for recalibration of the mass flow	
## Wetted and Seal Material: Recommended Body material and seals suitable for the gas (preferably S3302)  ## Accuracy: ± 1% of actual value or (0.8% of Reading ± 0.2% of Full Scale)  ## Pressure sensitivity: ± 0.01% of full scale / PSI  ## It should have rigid metallic construction  ## It should have rigid metallic construction  ## It should be operated in the power supply of 230 V AC, 50 Hz  ## It should be operated in the power supply of 230 V AC, 50 Hz  ## It should be operated in the power supply of 230 V AC, 50 Hz  ## It should be operated in the power supply of 230 V AC, 50 Hz  ## It should have the connection of compression gas fittings  ## It should be operated in the power supply of 230 V AC, 50 Hz  ## It should be operated in the power supply of 230 V AC, 50 Hz  ## It should be possible by interfacing through the in build keypad with the display on the MFC  ## It should Brown the in build keypad with the display on the MFC  ## It spical Response Time <100 ms (Adjustable)  ## It should Response Time <100 ms (Adjustable)  ## It should Connection Options: DB9 Pin or RS 232  ## Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy Configuration advantage data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  ## All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  ## Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  ## Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  ## Manufacturers should have their own Re-Calibration and Service	3	Operating Pressure: Max 160 PSIA	
suitable for the gas (preferably SS302)  6	4	Temperature range of operation: -10 ° C to 50 ° C	
7 Pressure sensitivity: ± 0.01% of full scale / PSI 8 It should have rigid metallic construction 9 The output of the controller should be 0 – 5 V or 4 - 20 mA signals 10 The controller should be provided with circuit protection 11 It should be operated in the power supply of 230 V AC, 50 Hz 12 The controller should have the connection of compression gas fittings 13 Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC 14 Typical Response Time <100 ms (Adjustable) 15 Control Range: 0.01% to 100% Full Scale 16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices 18 All connecting cables/chords/interfaces ports and necessary power subjective to 230V converter) should be supplied along with the instrument 19 Detailed service manual and operating manual with circuit diagram should be provided along with the instrument Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation and Service	5	· ·	
8  t should have rigid metallic construction 9  The output of the controller should be 0 – 5 V or 4 - 20 mA signals 10  The controller should be provided with circuit protection 11  t should be operated in the power supply of 230 V AC, 50 Hz 12  The controller should have the connection of compression gas fittings 13  Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC 14  Typical Response Time <100 ms (Adjustable) 15  Control Range : 0.01% to 100% Full Scale 16  Electrical Connection Options: DB9 Pin or RS 232	6		
The output of the controller should be 0 – 5 V or 4 - 20 mA signals  10 The controller should be provided with circuit protection  11 t should be operated in the power supply of 230 V AC, 50 Hz  12 The controller should have the connection of compression gas fittings  13 Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14 Typical Response Time <100 ms (Adjustable)  15 Control Range : 0.01% to 100% Full Scale  16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter ) should be supplied along with the instrument  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	7	Pressure sensitivity: ± 0.01% of full scale / PSI	
10 The controller should be provided with circuit protection  11 It should be operated in the power supply of 230 V AC, 50 Hz  12 The controller should have the connection of compression gas fittings  13 Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14 Typical Response Time <100 ms (Adjustable)  15 Control Range: 0.01% to 100% Full Scale  16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17 Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  19 Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	8	It should have rigid metallic construction	
11 It should be operated in the power supply of 230 V AC, 50 Hz  12 The controller should have the connection of compression gas fittings  13 Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14 Typical Response Time <100 ms (Adjustable)  15 Control Range: 0.01% to 100% Full Scale  16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17 Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  19 Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	9	The output of the controller should be 0 – 5 V or 4 - 20 mA signals	
The controller should have the connection of compression gas fittings  Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14 Typical Response Time <100 ms (Adjustable)  Control Range: 0.01% to 100% Full Scale  16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	10	The controller should be provided with circuit protection	
Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14  Typical Response Time <100 ms (Adjustable)  15  Control Range : 0.01% to 100% Full Scale  16  Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17  Configuration and experiment setup reloads, data capturing and logging (including a graphing tool),simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (100 to 230V converter ) should be supplied along with the instrument  19  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service		It should be operated in the power supply of 230 V AC, 50 Hz	
Setting Flow Ramp up programming cycle should be possible by interfacing through the in build keypad with the display on the MFC  14  Typical Response Time <100 ms (Adjustable)  15  Control Range : 0.01% to 100% Full Scale  16  Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17  Configuration and experiment setup reloads, data capturing and logging (including a graphing tool),simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter ) should be supplied along with the instrument  19  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service		The controller should have the connection of compression gas fittings	
15 Control Range: 0.01% to 100% Full Scale  16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17 Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  19 Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	13		
Control Range: 0.01% to 100% Full Scale  16 Electrical Connection Options: DB9 Pin or RS 232  Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter) should be supplied along with the instrument  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	14	Typical Response Time <100 ms (Adjustable)	
Software: A PC based software program that should interface with the RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  17 Configuration and experiment setup reloads, data capturing and logging (including a graphing tool),simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter ) should be supplied along with the instrument  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	15	Control Range : 0.01% to 100% Full Scale	
RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy  Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for multiple devices  All connecting cables/chords/interfaces ports and necessary power supply (110V to 230V converter ) should be supplied along with the instrument  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	16	Electrical Connection Options: DB9 Pin or RS 232	
supply (110V to 230V converter ) should be supplied along with the instrument  Detailed service manual and operating manual with circuit diagram should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	17	RS-232 and RS 485 communication. The graphical user interface (GUI) should provide automatic configuration, session saving for easy Configuration and experiment setup reloads, data capturing and logging (including a graphing tool), simple and advanced script building for automating meter sequences, software alarms and support for	
should be provided along with the instrument  Necessary accessories such as Power Supply, Communication cable, software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	18	supply (110V to 230V converter ) should be supplied along with the	
software should be provided for trouble free operation of the instrument.  Manufacturers should have their own Re-Calibration and Service	19	·	
21	20	software should be provided for trouble free operation of the	
	21		

ADDITIONAL TERMS AND CONDITIONS				
22	Warranty	1 year from the date of installation		
23		Installation, Commissioning, Training and Service should be done by the supplier		

Note: All the 6 Digital Mass Flow Controller(MFC) Should be from the same OEM.

(Note: It is mandatory for the bidders to provide the compliance statement in tabular column format along with catalogue page number (comply/not comply) for the above points with document proof as required. Failing which bidders will be technically disqualified)

SIGNATURE OF BIDDER ALONG WITH SEAL OF THE COMPANY WITH DATE

# FINANCIAL BID (PROFORMA) - BILL OF QUANTITIES (BOQ) Item Name: "Digital Mass Flow Controller" Tender No. AE/SUJI/06/IOE23/DMFCON

It. No	Description of work	Quantity	Units	Basic Rate in INR	GST in Percentage	Total Amount with taxes in INR
Digit	al Mass Flow Controller					
1	Digital mass flow controller (Flow Range to 2000 SLPM)	1	Nos.			
2	Digital mass flow controller (Flow Range to 1000 SLPM)	2	Nos.			
3	Digital mass flow controller (Flow Range to 500 SLPM)	2	Nos.			
4	Digital mass flow controller (Flow Range to 250 SLPM)	1	Nos.			
5	Digital mass flow controller (Flow Range to 100 SLPM)	2	Nos.			
6	Digital mass flow controller (Flow Range to 50 SLPM)	1	Nos.			
	Grand Total					

Total Amount Rupe	ees in words		
1		 	



# CENTRE FOR INDUSTRIAL CONSULTANCY & SPONSORED RESEARCH (IC&SR) INDIAN INSTITUTE OF TECHNOLOGY MADRAS CHENNAI 600 036



# SETTLEMENT (RTGS) FACILITY FOR RECEIVING PAYMENTS

## A. Details of Account Holder

Name of the Institution	Indian Institute of Technology - Madras
Complete Contact Address	Industrial Consultancy and Sponsored Research Indian Institute of Technology-Madras, IIT- Madras Campus Post Office, Sardar Patel Road, Guindy, CHENNAI - 600 036
Telephone No./ Fax No.	Tel - 044-2257 8356
E- mail ID of the FO/AO/REG/DIR	dricsr@iitm.ac.in

## B. Bank Account Details:

Institution Account Name (As per Bank Record)	The Registrar, Indian Institute of Technology - Madras
Account No.	2722101003872
Account Print Name	IIT F A/C , The Registrar IIT Madras
IFSC CODE	CNRB0002722
Bank Name (in full)	Canara Bank
Branch Name	IIT-Madras Branch
Complete Branch Address	Canara Bank, IIT-Madras Branch, IIT- Madras Campus Post Office, Sardar Patel Road, Guindy, CHENNAI - 600 036
MICR No.	600015085
Account Type	Savings Account

Certified that the Institute's account is in an RTGS enabled branch.

I hereby declare that the particulars given above are correct and complete,

Date:

Signature of the competent Authority of the Institution with seal.

# FORMAT FOR AFFIDAVIT OF SELF-CERTIFICATION UNDER PREFERENCE TO MAKE IN INDIA – PER ITEM

Tender Reference Number:	
Name of the item / Service:	
Date: S/o, D/o. Resident of	, W/o,
Hereby solemnly affirm and declare as under	er:
Policy vide GoI Order no. P-45021/2/2017 28.05.2018, 29.05.2019and 04.06.2020) M	nd conditions of the Public Procurement (Preference to Make in India) 7-PP (B.EII) dated 15.06.2017 (subsequently revised vide orders dated OCI order No. 45021/2/2017-PP (BE II) Dt.16th September 2020 & P-10) Dt.4th March 2021 and any subsequent modifications/Amendments,
That the local content for all inputs which responsible for the correctness of the claims	constitute the said item/service/work has been verified by me and I am s made therein.
that Local Content is equal to or category.  I/We	ame of the supplier] hereby confirm in respect of quoted items more than 50% and come under "Class-I Local Supplier"  name of the supplier] hereby confirm in respect of quoted items  where but less than 50% and come under "Class-II Local Supplier"  sich the local value addition is made and the proportionate value of local  Percentage of Local content:
For and on behalf of	(Name of firm/entity)
Authorized signatory (To be duly authorized  Insert Name, Designation and Contact No.	· · · · · · · · · · · · · · · · · · ·
statutory auditor or cost auditor of the com	n excess of Rs. 10 Crores, the bidders shall provide this certificate from pany (in the case of companies) or from a practicing cost accountant or ct of suppliers other than companies) giving the percentage of local
This letter should be on the letterhead of submission of this will lead to Disqualific	the quoting firm and should be signed by a competent authority. Non ation of bids.

## $\underline{Annexure-F}$

	(To be given on the letter head of the	he bidder)
No		Dated:
	<u>CERTIFICATE</u>	
	(Bidders from India)	
_	ng restrictions on procurement from certify that I am not from such a cour	a bidder of a country which shares a land ntry.
	OR (whichever is applicabl	le)
(Bidde	ers from Country which shares a land	border with India)
border with India and hereby registered with the Competen	certify that I from It Authority. I also certify that I fulfil	a bidder of a country which shares a land (Name of Country) and has been all the requirements in this regard and is ne Competent Authority is to be attached)
Place: Date:		Signature of the Tenderer Name & Address of the Tenderer with Office Stamp