

#### INDIAN INSTITUTE OF TECHNOLOGY MADRAS Chennai 600 036

Telephone: [044] 2257 9763 E-mail: tender@imail.iitm.ac.in



Date:21.03.2023

The Manager (Project Purchase)

Open Tender Reference No:SAIF/BHAT/90/IOE23/TQICPMSICLA

GEM NAR ID: GEM/GARPTS/20032023/YAY3Y5SEC1V0 Due Date/Time:10.04.2023@ 3:00 PM

Dear Sir/Madam,

On behalf of the Indian Institute of Technology Madras, Tenders are invited in two bid system from Class-I local suppliers and Class II local suppliers, for the supply of "Triple quad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) with Ion Chromatograph (IC) and Laser Ablation (LA) accessories" Conforming to the specifications given in Annexure -A.

Tender Documents may be downloaded from Central Public Procurement Portal <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a>. Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a>. The portal enrolment is free of cost. Bidders are advised to go through instructions provided at "Help for contractors". [Special Instructions to the Contractors/Bidders for the e-submission of the bids online through this eProcurement Portal"]

Bidders can access tender documents on the website (For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, click on "GO" button to view all IIT Madras tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a> as per the schedule attached.

| <u>1)</u> | Pre-bid Meeting<br>Details  | :  | NA  |
|-----------|-----------------------------|----|---|
| <u>2)</u> | ICSR Vendor<br>Registration | •• | Vendor registration code. Vendor registration with IC&SR (IITM) is mandatory for bidders to participate in tenders.  ** For Vendor Registration & Guidelines, Please follow the website: https://icandsr.iitm.ac.in/vendorportal; Helpdesk: vendorhelpdesk@icsrpis.iitm.ac.in |

<u>No manual bids will be accepted.</u> All tender documents including Technical and Financial bids should be submitted in the E-procurement portal.

| Last date for receipt of tender  |   | 10.04.2023 @ 3:00 PM |
|----------------------------------|---|----------------------|
| Date & time of opening of tender | : | 11.04.2023 @ 3:00 PM |

### 3. Instructions to the Bidder:

| _         | ı                                 |   |   |  |
|-----------|-----------------------------------|---|---|--|
| <u>A)</u> | A) Searching for tender documents |   | <ul> <li>There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, organization name, location, date, value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.</li> <li>Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective "My Tender" folder. This would enable the CPP Portal to intimate the bidders through SMS / email in case there is any corrigendum issued to the tender document.</li> </ul>  |  |
|           |                                   |   | • The bidder should make a note of the <b>unique Tender ID</b> assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.   |  |
| <u>B)</u> | Assistance to bidders             | : | <ul> <li>Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.</li> <li>Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is [0120-4200462, 0120-4001002, 0120-4001005]</li> </ul>  |  |
| <u>C)</u> | <b>Enrollment Process</b>         | : | REGISTRATION  |  |
|           | to Bidders                        |   | <ul> <li>Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal URL:https://etenders.gov.in/eprocure/app by clicking on "Online Bidder Enrollment". Enrollment on the CPP Portal is free of charge.</li> <li>As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.</li> <li>Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.</li> <li>Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.)</li> <li>Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse.</li> <li>Bidder then may log in to the site through the secured log-in by entering their user ID / password and the password of the DSC / eToken.</li> <li>Possession of a Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/e-token in the company's name</li> </ul> |  |

|           |                     |   | is a prerequisite for registration and participating in the bid submission activities through https://etenders.gov.in/eprocure/app  • Digital Signature Certificates can be obtained from the authorized certifying agencies, details of which are available in the web site https://etenders.gov.in/eprocure/app under the "Information about DSC".   |
|-----------|---------------------|---|--|
| <u>D)</u> | Preparation of bids | : | Bidder should take into account any corrigendum published on the tender document before submitting their bids.   |
|           |                     |   | <ul> <li>Please go through the tender advertisement and the tender<br/>document carefully to understand the documents required to be<br/>submitted as part of the bid. Please note the number of covers in<br/>which the bid documents have to be submitted, the number of<br/>documents including the names and content of each of the<br/>document that need to be submitted. Any deviations from these may<br/>lead to rejection of the bid.</li> </ul>   |
|           |                     |   | <ul> <li>Bidder, in advance, should prepare the bid documents to be<br/>submitted as indicated in the tender document / schedule and<br/>generally shall be in PDF / XLS formats as the case may be. Bid<br/>documents may be scanned with 100 dpi with black and white<br/>option.</li> </ul>   |
|           |                     |   | • To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, GSTIN Details, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use "My Documents" area available to them to upload such documents. These documents may be directly submitted from the "My Documents" area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process. |
| <u>E)</u> | Submission of bids  | : | Bidder should log into the site well in advance for bid submission so that he/she can upload the bid in time i.e. on or before the bid submission date and time. Bidder will be responsible for any delay due to other issues.   |
|           |                     |   | The bidder has to digitally sign and upload the required bid documents one by one as indicated in the tender document.   |
|           |                     |   | <ul> <li>Bidder has to select the bid security declaration. Otherwise, the<br/>tender will be summarily rejected.</li> </ul>   |
|           |                     |   | • A standard BOQ format has been provided with the tender<br>document to be filled by all the bidders. Bidders are requested to<br>note that they should necessarily submit their financial bids in the<br>format provided and no other format is acceptable. Bidders are<br>required to download the BOQ file, open it and complete the detail<br>with their respective financial quotes and other details (such as<br>name of the bidder). If the BOQ file is found to be modified by the<br>bidder, the bid will be rejected.   |
|           |                     |   | • The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.  |

|           |                             | • The Tender Inviting Authority (TIA) will not be held responsible for any sort of delay or the difficulties faced during the submission of bids online by the bidders due to local issues.   |
|-----------|-----------------------------|---|
|           |                             | <ul> <li>The uploaded tender documents become readable only after the<br/>tender opening by the authorized bid openers.</li> </ul>  |
|           |                             | • Upon the successful and timely submission of bids, the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.  |
|           |                             | <ul> <li>Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.</li> <li>More information useful for submitting online bids on the CPP Portal may be obtained at: <a href="https://etenders.gov.in/eprocure/app">https://etenders.gov.in/eprocure/app</a>.</li> <li>All tender documents including pre-qualification bid, Technical Bid &amp;Financial Bid should be submitted separately in online CPP portal as per the specified format only. Right is reserved to ignore any tender which fails to comply with the above instructions. No manual bid submission will be entertained.</li> </ul> |
| <u>F)</u> | Marking on<br>Technical Bid | • The bidder eligibility criteria, technical specification and supply of item for this tender is given in Annexure A.   |
|           |                             | • The Bidders shall go through the specification and submit the technical bid.  |
|           |                             | • The Technical bid should be submitted in the proforma as per Annexure-B in pdf format only through online (e-tender). No manual submission of bid will be entertained.  |
|           |                             | • The technical bid should have a page-wise heading as "Technical Bid" and page no. in all pages with seal and signature of authorized signatory. The total no. of pages should be mentioned at the last page of the documents.   |
|           |                             | • The technical bid should consist of bidder eligibility criteria details and all technical details along with catalogue/ pamphlet which will give a detailed description of product with technical data sheet so that technical compliance can be verified.  |
| <u>G)</u> | Marking on Price Bid        | • Financial bid (BoQ) should be submitted in the prescribed proforma format as per Annexure-C in xls format through e-tender only. No manual or other form of submission of Financial Bid will not be entertained   |
| L         | 1                           |   |

4) **Preparation of Tender**: The bidders should submit the bids in two bid system as detailed below.

#### Bid I \_Technical Bid

The technical bid should consist of bidder eligibility criteria and technical specification compliance sheet as per Annexure-B.

#### Bid II \_Price Bid

The price bid should be submitted in excel format (BoQ) as per the proforma (Annexure C) uploaded in the e-Tender web site. The Quoted price should be for supply and installation of the item and inclusive of all cost and statutory levies at IIT Madras.

#### 5) Price:

- a) The price should be quoted only in INR net per unit (after breakup) and must include all packing, transit insurance and delivery charges to the entry of **Sophisticated Analytical Instruments Facility.**
- b) The rate quoted shall be all inclusive of all taxes and no extra payment will be made other than statutory revisions as per the terms and conditions stipulated in this contract document.
- c) The percentage of tax & duties should be clearly indicated separately. IIT Madras is eligible for custom duty (5.5%). Relevant certificates will be issued wherever necessary.
- d) The offer/bids should be submitted through online only in two bid system i.e. Technical Bid and Financial Bid separately.

#### 6) Tenderer shall submit along with this tender:

- (i) Proof of having ISO or other equivalent certification given by appropriate authorities.
- (ii) Name and full address of the Banker and their swift code and PAN No. and GSTIN number.
- (iii) GST registration proof showing registration number, area of registration etc.
- (iv) All of your future correspondences including Invoices should bear the GST No. and Area Code.

#### 7) Terms of Delivery:

Supplier will be fully responsible for the safe carriage, Installation/Commissioning of goods up to the **Sophisticated Analytical Instruments Facility**, IIT Madras or named place as per PO, Insurance coverage will be in the scope of the supplier.

The tenderer should indicate clearly the time required for delivery of the item (subject to the approval of the Executive Committee-IIT-Madras). In case there is any deviation in the delivery schedule, liquidated damages clause will be enforced or penalty for the delayed supply period will be levied.

In the event of delay or non-supply of materials/execution of Contract beyond the date of delivery/completion of job. The penalty will be levied @1% per week of delay subject to a max of 10% of the value of purchase order and if the delay is more than accepted time frame by IIT M, the PO would be partially or fully cancelled and liquidated damages will be enforced accordingly.

#### 8) Period for which the offer will remain open:

The Tender shall remain open for acceptance/validity till: 120 days from the date of opening of the tender. However, the day up to which the offer is to remain open being declared closed holiday for the Indian Institute of Technology Madras, the offer shall remain open for acceptance till the next working day.

#### 9) **EMD**:

The EMD of **Rs.7,74,000** to be transferred to the account details mentioned in Annexure D and proof should be enclosed in the Technical Bid. Any offer not accompanied with the EMD shall be rejected summarily as non-responsive.

The EMD of the unsuccessful bidders shall be returned within 30 days of the end of the bid validity period. The same shall be forfeited, if the tenderers withdraw their offer after the opening during the bid validity period. The Institute shall not be liable for payment of any interest on EMD.

EMD is exempted for Micro and Small Enterprises (MSE) as defined in MSE Procurement Policy issued by Department of Micro, Small and Medium Enterprises (MSME) and Startups as recognized by Department of Industrial Policy & Promotion (DIPP). (MSE/MSME/DIPP PROOF should be enclosed in the cover containing technical bid)

#### **10**) **Performance Security: -**The successful bidder should submit Performance Security for an amount of 3% of the value of the contract/supply. The Performance Security may be furnished in the form of an Account Payee DD, FD Receipt in the name of "The Registrar, IIT Madras" from any scheduled commercial bank or Bank Guarantee from any scheduled commercial bank in India. The performance security should be furnished within 14 days from the date of the purchase order. Performance Security in the form of Bank Guarantee: - In case the successful bidder wishes to submit Performance Security in the form of Bank Guarantee, the Bank Guarantee should be routed directly to IIT Madras from the Bank. The Bank Guarantee should remain valid for a period of sixty days beyond the date of completion of all contractual obligations of the supplier including the warranty obligations. For the same tender, either the OEM or the authorized dealer/service provider can only quote. But both 11) of them cannot quote separately for the same tender. The offers/bids should be sent only for a item/Equipments of latest version that is available in the market **12**) and supplied to a number of customers. A list of customers in India with details must accompany the quotations. Quotations for a prototype machine will not be accepted 13) Original catalogue (not any photocopy) of the quoted model duly signed by the principals must accompany the quotation in the Technical bid. Compliance or Confirmation report with reference to the specifications and other terms & conditions **14**) should also be obtained from the principal/OEM. **15**) **Risk Purchase Clause** In the event of failure of supply of the item/equipment within the stipulated delivery schedule, the purchaser has all the right to purchase the item/equipment from other sources on the total risk of the supplier under risk purchase clause. **16**) **Payment:** No Advance payment will be made. However, 90% Payment against Delivery and 10% after installation are agreed to wherever the installation is involved. Advance Payment: No advance payment is generally admissible. In case a specific percentage of advance payment is required, the Vendor has to submit a Bank Guarantee from a scheduled commercial bank in India equivalent to the amount of advance payment. **17**) **On-site Installation:** The equipment/item or Machinery has to be installed or commissioned by the successful bidder within the number of days (as prescribed by PI) from the date of receipt of the item at the site of IIT Madras. **18**) Warranty/Guarantee: The offer should clearly specify the warranty or guarantee period for the machinery/equipment. Any extended warranty offered for the same has to be mentioned separately (For more details please refer our Technical Specifications). \*\* Note: PO which involves installation, warranty/guarantee shall be applicable from date of installation. **19**) **Acceptance and Rejection:** Failure to comply with any of the instructions stated in this document or offering unsatisfactory explanations for non-compliance will likely to lead to rejection of offers.

I.I.T. Madras has the right to accept the whole or any part of the Tender or portion of the

quantity offered or reject it in full without assigning any reason.

#### **20)** Debarment from Bidding:

In case of breach of Terms & Conditions, Bidder may be suspended from being eligible for bidding in any contract with the IIT Madras up to 2 Years [as per Rule 151(iii) of GFR] from the date of Tender.

### 21) Disputes and Jurisdiction:

Settlement of Disputes: Any dispute, controversy or claim arising out of or in connection with this PO including any question regarding its existence, validity, breach or termination, shall in the first instance be attempted to be resolved amicably by both the Parties. If attempts for such amicable resolution fails or no decision is reached within 30 days whichever is earlier, then such disputes shall be settled by arbitration in accordance with the Arbitration and Conciliation Act, 1996. Unless the Parties agree on a sole arbitrator, within 30 days from the receipt of a written request by one Party from the other Party to so agree, the arbitral panel shall comprise of three arbitrators. In that event, the supplier will nominate one arbitrator and the Project Coordinator of IITM shall nominate on arbitrator. The Dean IC&SR will nominate the Presiding Arbitrator of the arbitral tribunal. The arbitration proceeding shall be carried out in English language. The cost of arbitration and fees of the arbitrator(s) shall be shared equally by the Parties. The seat of arbitration shall be at IC&SR IIT Madras, Chennai.

- a. **The Applicable Law:** The Purchase Order shall be construed, interpreted and governed by the Laws of India. Court at Chennai shall have exclusive jurisdiction subject to the arbitration clause.
- b. Any legal disputes arising out of any breach of contact pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Chennai in Tamil Nadu.
- **Force Majeure:** The Supplier shall not be liable for forfeiture of its performance security, liquidated damages or termination for default, if and to the extent that, it's delay in performance or other failure to perform its obligations under the Contract is the result of an event of Force Majeure.

For purposes of this Clause, "Force Majeure" means an event beyond the control of the Supplier and not involving the Supplier's fault or negligence and not foreseeable. Such events may include, but are not limited to, acts of the Purchaser either in its sovereign or contractual capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.

If a Force Majeure situation arises, the Supplier shall promptly notify the Purchaser in writing of such conditions and the cause thereof. Unless otherwise directed by the Purchaser in writing, the Supplier shall continue to perform its obligations under the Contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the Force Majeure event.

## 23) Eligibility Criteria:

- > As per the Government of India Order, only "Class I Local Suppliers" and "Class II Local Suppliers" can participate in this tender.
- Bidder should confirm their acceptance that they comply with the provisions with report to "Guidelines for eligibility of a bidder from a country which shares a land border with India as detailed at Annexure-F. The bidder should submit Certificate for "Bidder from/
  Not from Country sharing Land border with India & Registration of Bidder with Competent Authority" as per Order of DoE F.No.6/18/2019-PPD dated 23.07.2020 as mentioned.
- Preference to "class I Local Suppliers": preference will be given to "class 1 local suppliers" (subject to class -I local supplier's quoted price falling within the margin of purchase preference ) as per public procurement (preference to make in India) order 2017 .O.M No P- 45021/2/2017 pp(BE 11) dt 04/06/2020 subject to the conditions that the "class 1 Local Supplier" should agree to supply goods / provide service at L1 rate and furnish a certificate with the technical bid document that the goods/service provided by them consists local content equal to or more than 50%.( certificate from

Chartered Accountant in case value of contract exceeds Rs 10 crore).

- > 'Class I local supplier' means a supplier or service provider whose goods, services or works offered for procurement consists of local content equal to or more than 50% as defined under the above said order. Declaration to be provided as per Annexure-D per item/service/work.
- > 'Class II local supplier' means a supplier or service provider whose goods, services or works offered for procurement consists of local content equal to 20% but less than 50% as defined under the above said order. Declaration to be provided as per Annexure-D per item/service/work.
- → 'Margin of purchase preference': The margin of purchase preference shall be 20%. The Definition of the margin of purchase preference is defined in the Govt. of India Order No: P-45021/12/2017-PP (BE-II) Dt.4th June, 2020) Order 2017. As per the Government of India Order "Margin of Purchase Preference" means the maximum extent to which the price quoted by a "Class-I local supplier" may be above the L1 for the purpose of purchase preference.

\*\*Note: Local content percentage to be calculated in accordance with the definition provided at clause 2 of revised public procurement preference to Make in India Policy vide GoI Order no. P-45021/2/2017-PP (B.E.-II) dated 15.06.2017 (subsequently revised vide orders dated 28.05.2018, 29.05.2019and 04.06.2020) MOCI order No. 45021/2/2017-PP (BE II) Dt.16th September 2020 & P-45021/102/2019-BE-II-Part(1) (E-50310) Dt.4th March 2021

### **Evaluation of Bids**

Bid evaluation will take place in two stages.

#### Stage I Technical Bid evaluation

All bidders who have fully complied with bidder eligibility criteria I, II and technical evaluation (Annexure A) will only be considered for opening of price bid.

#### **Stage II: Price Bid Evaluation**

The price bid evaluation will be based on price quoted by the bidder. The rate quoted for "Triple quad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) with Ion Chromatograph (IC) and Laser Ablation (LA) accessories" unit will alone be taken up for arrival of Lowest Bid (L1) value.

- 26) Selection of successful bidder and Award of Order
  - The order will be directly awarded to the technically qualified bidder as per the condition in para 3A of DIPP, MoCI Order No. 45021/2/2017-PP (BE II) dated 16th September 2020.
- All information including selection and rejection of technical or financial bids of the prospective bidders will be communicated through e-Tender portal. In terms of Rule 173(iv) of General Financial Rule 2017, the bidder shall be at liberty to question the bidding conditions, bidding process and/or rejection of bids.
- The tenderer shall certify that the tender document submitted by him / her are of the same replica of the tender document as published by IIT Madras and no corrections, additions and alterations made to the same. If any deviation found in the same at any stage and date, the bid / contract will be rejected / terminated and actions will be initiated as per the terms and conditions of the contract.
- Due to Covid-19 pandemic pre-bid meeting will be conducted through online. Clarification to the queries and doubts raised by the bidders will be issued as a corrigendum/addendum in the e-tenders portal.
- Due to Covid-19 pandemic the bidders will not be entertained to participate in opening of Bids. Since the tender is e-tender, the opening of the bids may be checked using the respective logins of the bidders.

#### **ACKNOWLEDGEMENT**

It is hereby acknowledged that I/We have gone through all the points listed under "Specification, Guidelines, Terms and Conditions" of tender document. I/We totally understand the terms and conditions and agree to abide by the same.

SIGNATURE OF TENDERER ALONG WITH SEAL OF THE COMPANY WITH DATE

Bidder Eligibility Criteria and Technical Specification for "Triple quad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) with Ion Chromatograph (IC) and Laser Ablation (LA) accessories" Tender No. SAIF/BHAT/90/IOE23/TQICPMSICLA

#### Bidder Eligibility Criteria – I (Public Procurement – Preference to Make in India)

Only 'Class-I local suppliers' and 'Class-II local suppliers', as defined under DIPP, MoCI Order No. P-45021/2/2017-PP (BE-II) dated 16<sup>th</sup> September 2020 and other subsequent orders issued therein.

#### Bidder Eligibility Criteria – II

1. The bidder/OEM should have supplied at least 3 similar equipments (not mandatory as a set) to IITs, NITs, IISERs, CSIR Labs or other Govt. organizations in the last 5 years, PO copies or installation certificates along with contact details of end user need to be submitted as the proof of supply. IIT Madras reserves its right to verify the claims submitted by the bidder and the feedback from the previous customers will be part of technical evaluation.

# III. Technical Specification For "Triple quad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) with Ion Chromatograph (IC) and Laser Ablation (LA) accessories"

#### DESCRIPTION

State-of-the art Triplequad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) capable of performing multi element analysis at major, minor, trace and ultra-trace levels in single run for different matrices namely environmental, chemical, biological and geological samples etc. The Ion chromatograph (IC) should have coupling option with the ICP-MS/MS and also be capable of functioning in standalone mode. Laser ablation (LA) accessory should be fully compatible with the ICP-MS/MS to obtain high end results for solid samples.

| S.No   | SPECIFICATION              |  |  |  |  |
|--|----------------------------|--|--|--|--|
| 1  | ICP-MS/MS                  |  |  |  |  |
|  | Sample Introduction system | Sample Introduction system compatible for aqueous, organic, samples with high TDS and HF based samples to be provided. Peltier cooled spray chamber (-5to 20 deg or better),inert torch tube, PFA/Glass/Quartz nebulizer, Ni sample and skimmer cones, Platinum sample and skimmer cones, Low pulsation, high precision multichannel peristaltic pump. Separate organic kit, HF kit, tubing set and MFC for oxygen to aspirate organic samples directly. Online internal standard kit with splitter and tubing set. Factory integrated dilution accessories with 100-fold or better dilution capabilities (UHMI/AMS/AGD) for direct aspiration of samples containing 25% TDS. Accessories must be fully automated and software controlled without any manual intervention necessary and without utilizing Internal Standard kit or tubing's. |  |  |  |
|  | RF generator and torch     | RF generator 27 MHz or more, 550-1600 W or better for automatic control of torch ignition, shutdown and system warm up. The plasma torch should have provision for software controlled alignment of horizontal, vertical position and sampling depth   |  |  |  |
|  | Ion extraction Interface   | Standard large orifice Ni sampler ( $1\pm0.1$ mm ID) and skimmer ( $0.45\pm0.05$ mm ID) cones with suitable diameters  |  |  |  |
|  | Ion focusing system        | The ion deflection system should have efficient mechanism for removing all neutrals from the Ion path. The ion path must be maintenance free and at a 90 deg deflection or off-axis or suitably designed   |  |  |  |
| The system should incorporate Standard mode, collision |                            | The system should incorporate Standard mode, collision mode (He) with KED and reaction mode for interference removal in a single analytical method   |  |  |  |

|   |   | simultaneously. The system should have collision and reaction cell (CRC) with controlled reaction capability to carry out mass shift reactions. Cell must be non-consumable with zero maintenance.   |  |  |
|---|---|--|--|--|
|   | Gas control   | The system should have factory fitted MFC (4 Nos) to control He (collision gas) and reaction gases like H <sub>2</sub> , NH <sub>3</sub> , O <sub>2</sub> etc in a safe premix or pure form and compliant manner. System should have provision to handle an additional of minimum 15 CRC gases based on request. System should have dedicated MFC devices to control Ar flow for plasma, auxiliary and nebulizer and Ar + O2 flow for samples containing high TDS and O2 flow for organic samples.   |  |  |
|   | Quadrupole assembly (for MS/MS<br>Triple quadrupole system)                               | Quadrupole mass analyzer Q1&Q3 must have unit mass (1 amu) resolution capability. User definable resolution from $\leq 0.3u$ to $\leq 1.0$ u should be possible. Q2 should be Collision – Reaction based. "Tandem mass spectrometer comprising two transmission quadrupole mass spectrometers in series, with a (non-selecting) RF-only quadrupole (or other multipoles) between them to act as a collision cell" as defined by the International Union of Pure and Applied Chemistry (IUPAC)) Mass calibration assessed and automatically updated. Mass range 4-260 u or better Scan speed >3700 amu/sec or better at 40 mass intervals . |  |  |
|   | Ion detector assembly   | Simultaneous dual mode discrete dynode electron multiplier. Dynamic range: 10 orders or more, without any Software/hardware adjustment Minimum dwell time of 100 µs or better in pulse and analog  |  |  |
|   | Vacuum system   | Suitable corrosion resistant turbo pump and other backing pumps should be supplied. Vacuum level 1x10 <sup>-5</sup> Torr or better.  |  |  |
|   | Performance specifications  | Sensitivity (Millicounts Per Second/Parts Per Million) Low mass Li/Be: ≥ 65 or better Mid mass In/Y: ≥ 300 or better High mass U/Tl:≥ 330or better  Detection limit no gas mode (Parts Per Trillion) Low mass Be/Li ≤ 0.5 or better Mid mass In/Y: ≤ 0.1 or better High mass Tl/U: ≤ 0.1 or better Oxide ratio(% RSD) CeO / Ce < 2% or better Back ground (No gas/KED mode) for Low and High mass < 1 counts per second or better Short term & long term stability (% RSD) < 4% or better Isotope ratio precision (% RSD): Ag 107/ Ag 109 ≤ 0.3% or better   |  |  |
|   | NPS module  | Single nanoparticle comprehensive module for estimation of number of particles, particle concentration, particle size and distribution should be provided.   |  |  |
|   | Autosampler   | Suitable autosampler to be provided  |  |  |
| 2 | Ion Chromatograph   |  |  |  |
|   | Design aspects  | Standalone Ion chromatograph suitable for analyzing anions, cations, transition metal ions, organic acids, amines etc by suppressed or non-suppressed mode conductivity measurements and speciation of Fe, Hg, As, Cr etc by coupling with ICP-MS/MS   |  |  |
|   | Solvent delivery pump   | Quarternary/ternary gradient pump with built in degassing system type Serial dual piston pump, microprocessor controlled, constant stroke, variable speed, Chemically inert, metal free PEEK pump heads and flow path, compatible with aqueous eluents from pH 0-14 and reverse phase solvents Pressure range: 0-5000 psi Flow range: 0.000 – 10 mL/min with settable flow increments at 0.001 mL/min Flow precision <0.1% Flow accuracy < 0.1% Pressure ripple <1% Temperature range: 30 to 60 deg  |  |  |
|   | IC power cords  | To be provided   |  |  |
|   | Eluent organizer with four 2 L<br>PP/PTFE bottles   | To be provided   |  |  |
|   | Inline eluent filters   | To be provided   |  |  |
|   | DC, Dual Independent Temperature zones, 6 port PEEK manual injection valve, Standard bore | To be provided   |  |  |
|   | Detector  | Thermostated microflow cell block Measurement range: 0-15000 μS/cm or better Cell drive: 8 KHz square wave or 1.4 KHz sine wave Conductivity resolution: 0.005 nS/cm or better Linearity: 1 % or better Baseline noise: ± 0.1nS or better Data sample rate: 10 ms or better Temperature range 20°C to 50 erutarepmeT retteb ro C'100.0 ±: ytilibats erutarepmeT elbattes resu ,C°  |  |  |

|     |   | compensation: 0-3% per °C Temperature compensation: 0-3% per °C Cell electrodes: compatible with majority of the used eluents Detector should be upgradable in future. The accessory for connection to ICP-MS detector for automatic triggering of data acquisition should be seamless integration (both ICP-MS/MS& IC).  |
|-----|---|---|
|     | Columns for Speciation analysis of Cr,<br>As, Hg and Fe by interface with ICP-<br>MS/MS | The following columns to be provided Analytical PEEK Column Guard PEEK Column   |
|     | Suppressor  | Suppressor should be of packed bed type with high loading and high back pressure (300 psi or more) tolerance with continuous regeneration. External chemical regeneration mode with controlled flow rate up to 10 mL/min. Should be 100 % solvent compatible.   |
| 3   | Laser ablation system   |   |
|     | Laser   | Frequency quintupled Nd:YAG laser 213 nm for good coupling with nearly all sample types, including highly transmitting materials.  Energy: 3 mJ/pulse or higher Pulse width: 5 ns Energy output: 0-100% Spot sizes: 4 µm to 200 µm dia must be achievable with aperture imaging, with larger spots achievable via defocusing Sampling repeatition rate: 1-20 Hz The number of laser transfer optics must be as low as possible, minimizing energy losses.   |
|     | CCD camera / Microscope   | Spot selection must be via ceramic aperture wheel.  Resolution: 2 μm or better  Field of view ≥ 6 mm of sample surface  Optical zoom 2.5x to 32x or better  |
|     | Lighting system   | Software controlled high intensity LED for sample illumination (Incident &transmitted) and cross polariser  |
|     | Sample cell / chamber   | Fast washout HelEx II or equivalent cell to be provided Capable of measuring samples upto 50mm or more in height X-Y stage resolution 0.16 µm or lower Stage travel 100 mm in the X and Y planes The washout performance of the chamber must be <1.5 s for washout to 1 % at any point in the cell area A sample cell gas purge system controlled by precision calibrated MFCs that uses Helium to transport and entrain the ablation particles into the argon stream of the ICP. Open architecture for maximum stability.  |
| 4   | Accessories for entire system (I  | CP-MS/MS + IC+ LA)  |
|     | Other items   | a) Latest PC's/workstations with 24" monitor and i7 OS compatible with ICP-MS/MS, IC and LA applications. The main system should have inbuilt hardware features required for seamless integration of Ion Chromatograph and Laser ablation accessories. An original licensed copy of the windows-based software which includes all features required for ICP-MS/MS operation, NPS module, Inorganic speciation and Laser ablation studies to be provided. b) Suitable noise free recirculating chiller should be provided. c) Exhaust system for ICP-MS/MS should be provided. d) Gas purification panel, manifold double stage for Argon cylinder, manifold single stage for He, H2, O2 and NH3 cylinders and gas regulators for all the gases mentioned above to be provided. e) Vibration free bench for housing ICP-MS/MS, IC and LA to be provided. |
|     | UPS   | High performance online UPS with in-built isolation transformer compatible for all the three modules namely ICP-MS/MS, IC and LA operating simultaneously with battery backup for 1-hour (at full load) to be provided  |
| ОТН | ER TERMS AND CONDITION Warranty   | Three years standard warranty from the date of installation/acceptance, covering  |

|      | all the units / accessories of the instrument to be provided.   |  |  |  |  |
|------|---|--|--|--|--|
| OPTI | OPTIONAL ITEMS  |  |  |  |  |
| SNO  | SPECIFICATION   |  |  |  |  |
| 1    | Additional comprehensive warranty for two years for all parts and peripherals to be quoted separately.  |  |  |  |  |
| 2    | Suitable auto sampler for IC and LA to be quoted separately.  |  |  |  |  |
| 3    | Suitable PEEK analytical and guard columns for determination of anions and cations to be quoted separately.   |  |  |  |  |
| 4    | Suitable PEEK analytical and guard columns for determination of transition metal ions to be quoted separately.  |  |  |  |  |
| 5    | Suitable column oven to be quoted separately.   |  |  |  |  |
| 6    | Suitable detector for determination of transition metal ions to be quoted separately.   |  |  |  |  |
| GEN  | ERAL TERMS AND CONDITIONS   |  |  |  |  |
| 1    | Price for each item/accessory of the instrument should be quoted separately. However, the evaluation for arriving L1 on LCS will be done based on the overall value of all major items given in Sl.Nos. I, II III and IV. |  |  |  |  |
| 2    | Certified calibration standards and solvents required for smooth operation of ICP-MSMS, IC and LA for the first five  |  |  |  |  |
|      | years should be provided along with the instrument.  a) Suitable  |  |  |  |  |
|      | aqueous standards (individual/ multielement/internal ) and organometallic standards in suitable organic solvents for ICP-   |  |  |  |  |
|      | MSMS applications should be provided.  b) Standards, cationic and anionic buffers and   |  |  |  |  |
|      | mobile phase modifiers required for speciation of Fe, Cr, Hg and Assuing IC should be provided. c) Glass  |  |  |  |  |
|      | standards (610, 612, 614, 616, MACS -3) and other standards (Be,Mg,P,S,V,Ge,Br, Y,Zr,Rh,Pd,Sn,Cs,Ta,W,Pt,Hg,Bi and  |  |  |  |  |
| 2    | Pr) for LA applications should be provided.   |  |  |  |  |
| 3    | Software upgrades should be incorporated by the vendor as and when the new versions are released at no additional cost.   |  |  |  |  |
| 4    | Spares and consumables needed for smooth operation of ICP-MSMS, IC and LA for the first five years to be provided   |  |  |  |  |
| 5    | along with the instrument.  |  |  |  |  |
|      | Accessories such as UPS and Chiller may be supplied locally.  |  |  |  |  |
| 6    | One week onsite training for the operation and general maintenance must be provided.  |  |  |  |  |

#### TECHNICAL BID PROFORMA

#### Tender No. SAIF/BHAT/90/IOE23/TQICPMSICLA

Item Name: "Triple quad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) with Ion Chromatograph (IC) and Laser Ablation (LA) accessories"

**Bidder Eligibility Criteria:** 

| 1.0 | Bidder Eligibility Criteria-I<br>(Public Procurement – Preference to Make in India)  | Class I /<br>Class II  | Local<br>Content<br>value | Reference,<br>Page No. |
|-----|--|------------------------|---------------------------|------------------------|
| I   | Only 'Class-I local suppliers' and 'Class-II local suppliers', as defined under DIPP, MoCI Order No. P-45021/2/2017-PP (BE II) dated 16 <sup>th</sup> September 2020 and other subsequent orders issued therein.   |                        |                           |                        |
| 2.0 | Bidder Eligibility Criteria-II   | Compliance<br>(Yes/No) | Reference<br>Page No.     | Remarks, If any        |
| 1   | The bidder/OEM should have supplied at least 3 similar equipments (not mandatory as a set) to IITs, NITs, IISERs, CSIR Labs or other Govt. organizations in the last 5 years, PO copies or installation certificates along with contact details of end user need to be submitted as the proof of supply. IIT Madras reserves its right to verify the claims submitted by the bidder and the feedback from the previous customers will be part of technical evaluation. |                        |                           |                        |

#### 3.0 Technical Compliance:

#### **DESCRIPTION**

State-of-the art Triplequad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) capable of performing multi element analysis at major, minor, trace and ultra-trace levels in single run for different matrices namely environmental, chemical, biological and geological samples etc. The Ion chromatograph (IC) should have coupling option with the ICP-MS/MS and also be capable of functioning in standalone mode. Laser ablation (LA) accessory should be fully compatible with the ICP-MS/MS to obtain high end results for solid samples.

| S.No |                                  | SPECIFICATION  | COMPLIED/NOT | REF. PG.NO |
|------|----------------------------------|--|--------------|------------|
|      |                                  |  | COMPLIED     |            |
| 1    | ICP-MS/MS                        |  |              |            |
|      | Sample<br>Introduction<br>system | Sample Introduction system compatible for aqueous, organic, samples with high TDS and HF based samples to be provided. Peltier cooled spray chamber (-5to 20 deg or better),inert torch tube, PFA/Glass/Quartz nebulizer, Ni sample and skimmer cones, Platinum sample and skimmer cones, Low pulsation, high precision multichannel peristaltic pump. Separate organic kit, HF kit, tubing set and MFC for oxygen to aspirate organic samples directly. Online internal standard kit with splitter and tubing set. Factory integrated dilution accessories with 100-fold or better dilution capabilities (UHMI/AMS/AGD) for direct aspiration of samples containing 25% TDS. Accessories must be fully automated and software controlled without any manual intervention necessary and without utilizing Internal Standard kit or tubing's. |              |            |

| P.E  | DE . 27 MI  | T T |
|--|---|-----|
| RF generator and torch   | RF generator 27 MHz or more, 550-1600 W or better for automatic control of torch ignition, shutdown and system warm up. The plasma torch should have provision for software controlled alignment of horizontal, vertical position and sampling depth  |     |
| Ion extraction<br>Interface  | Standard large orifice Ni sampler (1±0.1 mm ID) and skimmer (0.45± 0.05 mm ID) cones with suitable diameters  |     |
| Ion focusing system  | The ion deflection system should have efficient mechanism for removing all neutrals from the Ion path. The ion path must be maintenance free and at a 90 deg deflection or off-axis or suitably designed  |     |
| Cell Technology  | The system should incorporate Standard mode, collision mode (He) with KED and reaction mode for interference removal in a single analytical method simultaneously. The system should have collision and reaction cell (CRC) with controlled reaction capability to carry out mass shift reactions. Cell must be non-consumable with zero maintenance.   |     |
| Gas control  | The system should have factory fitted MFC (4 Nos) to control He (collision gas) and reaction gases like H <sub>2</sub> , NH <sub>3</sub> , O <sub>2</sub> etc in a safe premix or pure form and compliant manner. System should have provision to handle an additional of minimum 15 CRC gases based on request. System should have dedicated MFC devices to control Ar flow for plasma, auxiliary and nebulizer and Ar + O2 flow for samples containing high TDS and O2 flow for organic samples.  |     |
| Quadrupole<br>assembly (for<br>MS/MS Triple<br>quadrupole<br>system) | Quadrupole mass analyzer Q1&Q3 must have unit mass (1 amu) resolution capability. User definable resolution from $\leq$ 0.3u to $\leq$ 1.0 u should be possible. Q2 should be Collision – Reaction based. "Tandem mass spectrometer comprising two transmission quadrupole mass spectrometers in series, with a (non-selecting) RF-only quadrupole (or other multipoles) between them to act as a collision cell" as defined by the International Union of Pure and Applied Chemistry (IUPAC)) Mass calibration assessed and automatically updated. Mass range 4-260 u or better Scan speed >3700 amu/sec or better at 40 mass intervals. |     |
| Ion detector assembly  | Simultaneous dual mode discrete dynode electron multiplier.Dynamic range: 10 orders or more, without any Software/hardware adjustment Minimum dwell time of 100 µs or better in pulse and analog  |     |
| Vacuum system  | Suitable corrosion resistant turbo pump and other backing pumps should be supplied. Vacuum level 1x10 <sup>-5</sup> Torr or better.   |     |
| Performance specifications   | Sensitivity (Millicounts Per Second/Parts Per Million) Low mass Li/Be: $\geq 65$ or better Mid mass In/Y: $\geq$ 300 or better High mass U/Tl: $\geq$ 330or better High mass U/Tl: $\geq$ 330or better Detection limit no gas mode (Parts Per Trillion) Low mass Be/Li $\leq$ 0.5 or better Mid mass In/Y: $\leq$ 0.1 or better High mass Tl/U: $\leq$ 0.1 or better Oxide ratio( % RSD) CeO / Ce < 2 % or better Back ground (No gas/KED mode) for Low and High mass < 1 counts per second or better Short term & long term stability (% RSD) < 4 % or better Isotope ratio precision (% RSD): Ag 107/ Ag 109 ≤ 0.3 % or better          |     |

|   |   | Single nanoparticle comprehensive module for estimation of  |  |
|---|---|---|--|
|   |   | number of particles, particle concentration, particle size and  |  |
|   | distribution should be provided.                |   |  |
|   | Autosampler                                     | Suitable autosampler to be provided   |  |
| 2 | Ion Chromatograph                               |   |  |
|   |   | Standalone Ion chromatograph suitable for analyzing   |  |
|   | D   | anions, cations, transition metal ions, organic acids,  |  |
|   | Design aspects                                  | amines etc by suppressed or non-suppressed mode conductivity measurements and speciation of Fe, Hg, As,   |  |
|   |   | Cr etc by coupling with ICP-MS/MS   |  |
|   |   | Quarternary/ternary gradient pump with built in   |  |
|   |   | degassing system type Serial dual piston pump,  |  |
|   |   | microprocessor controlled, constant stroke, variable  |  |
|   |   | speed, Chemically inert, metal free PEEK pump heads   |  |
|   | Solvent delivery pump                           | and flow path, compatible with aqueous eluents from pH  |  |
|   |   | 0-14 and reverse phase solvents. Pressure range: 0-5000   |  |
|   |   | psi Flow range: 0.000 – 10 mL/min with settable flow increments at 0.001 mL/min Flow precision <0.1% Flow |  |
|   |   | accuracy < 0.1% Pressure ripple <1% Temperature range   |  |
|   |   | : 30 to 60 deg  |  |
|   | IC power cords                                  | To be provided  |  |
|   | Eluent organizer with                           | To be provided  |  |
|   | four 2 L PP/PTFE                                |   |  |
|   | bottles   |   |  |
|   | Inline eluent filters                           | To be provided  |  |
|   | DC, Dual Independent                            |   |  |
|   | Temperature zones, 6 port PEEK manual           | To be provided  |  |
|   | injection valve,                                | To be provided  |  |
|   | Standard bore                                   |   |  |
|   |   | Thermostated microflow cell block Measurement range:  |  |
|   |   | 0-15000 μS/cm or better Cell drive: 8 KHz square wave   |  |
|   |   | or 1.4 KHz sine wave Conductivity resolution: 0.005   |  |
|   |   | nS/cm or better Linearity: 1 % or better Baseline noise:  |  |
|   |   | ± 0.1nS or better Data sample rate : 10 ms or better<br>Temperature range 20°C to 50 °C, user settable    |  |
|   | Detector  | Temperature stability: $\pm 0.001$ °C or better Temperature   |  |
|   |   | compensation: 0-3% per °C Temperature compensation  |  |
|   |   | : 0-3% per °C Cell electrodes: compatible with majority   |  |
|   |   | of the used eluents Detector should be upgradable in  |  |
|   |   | future. The accessory for connection to ICP-MS detector   |  |
|   |   | for automatic triggering of data acquisition should be  |  |
|   | Columns for Smooistis                           | seamless integration (both ICP-MS/MS& IC).  |  |
|   | Columns for Speciatio<br>analysis of Cr, As, Hg |   |  |
|   | and Fe by interface                             | Guard PEEK Column   |  |
|   | with ICP-MS/MS                                  |   |  |
|   | Suppressor                                      | Suppressor should be of packed bed type with high   |  |
|   |   | loading and high back pressure (300 psi or more)  |  |
|   |   | tolerance with continuous regeneration. External  |  |
|   |   | chemical regeneration mode with controlled flow rate up to 10 mL/min. Should be 100 % solvent compatible. |  |
|   |   | •   |  |
| 3 | Laser ablation sys                              | stem  |  |

|     | Laser                                 | Frequency quintupled Nd:YAG laser 213 nm for good coupling with nearly all sample types, including highly transmitting materials.  Energy: 3 mJ/pulse or higher Pulse width: 5 ns Energy output: 0-100% Spot sizes: 4 µm to 200 µm dia must be achievable with aperture imaging, with larger spots achievable via defocusing Sampling repeatition rate: 1-20 Hz The number of laser transfer optics must be as low as possible, minimizing energy losses. Spot selection must be via ceramic aperture wheel.  |  |
|-----|---------------------------------------|---|--|
|     | CCD camera /<br>Microscope            | Resolution: $2 \mu m$ or better<br>Field of view $\geq 6 \text{ mm}$ of sample surface<br>Optical zoom 2.5x to 32x or better  |  |
|     | Lighting system                       | Software controlled high intensity LED for sample illumination (Incident &transmitted) and cross polariser  |  |
|     | Sample cell / chamber                 | Fast washout HelEx II or equivalent cell to be provided Capable of measuring samples upto 50mm or more in height X-Y stage resolution 0.16 µm or lower Stage travel 100 mm in the X and Y planes The washout performance of the chamber must be <1.5 s for washout to 1 % at any point in the cell area A sample cell gas purge system controlled by precision calibrated MFCs that uses Helium to transport and entrain the ablation particles into the argon stream of the ICP. Open architecture for maximum stability.  |  |
| 4   | Accessories for enti                  | re system (ICP-MS/MS + IC+ LA)  |  |
|     | Other items                           | a) Latest PC's/workstations with 24" monitor and i7 OS compatible with ICP-MS/MS, IC and LA applications. The main system should have inbuilt hardware features required for seamless integration of Ion Chromatograph and Laser ablation accessories. An original licensed copy of the windows-based software which includes all features required for ICP-MS/MS operation, NPS module, Inorganic speciation and Laser ablation studies to be provided. b) Suitable noise free recirculating chiller should be provided. c) Exhaust system for ICP-MS/MS should be provided. d) Gas purification panel, manifold double stage for Argon cylinder, manifold single stage for He, H2, O2 and NH3 cylinders and gas regulators for all the gases mentioned above to be provided. e) Vibration free bench for housing ICP-MS/MS, IC and LA to be provided. |  |
|     | UPS                                   | High performance online UPS with in-built isolation transformer compatible for all the three modules namely ICP-MS/MS, IC and LA operating simultaneously with battery backup for 1-hour (at full load) to be provided  |  |
| OTH | ER TERMS AND C                        | ONDITION  |  |
| 1   | Warranty                              | Three years standard warranty from the date of installation/acceptance, covering all the units / accessories of the instrument to be provided.  |  |
|     | · · · · · · · · · · · · · · · · · · · |   |  |

| OPTIONAL ITEMS |  |  |
|----------------|--|--|
| SNO            | SPECIFICATION  |  |
| 1              | Additional comprehensive warranty for two years for all parts and peripherals to be quoted separately.   |  |
| 2              | Suitable auto sampler for IC and LA to be quoted separately.   |  |
| 3              | Suitable PEEK analytical and guard columns for determination of anions and cations to be quoted separately.  |  |
| 4              | Suitable PEEK analytical and guard columns for determination of transition metal ions to be quoted separately.   |  |
| 5              | Suitable column oven to be quoted separately.  |  |
| 6              | Suitable detector for determination of transition metal ions to be quoted separately.  |  |
| GENI           | ERAL TERMS AND CONDITIONS  |  |
| 1              | Price for each item/accessory of the instrument should be quoted separately.  However, the evaluation for arriving L1 on LCS will be done based on the overall value of all major items given in Sl.Nos. I, II III and IV.   |  |
| 2              | Certified calibration standards and solvents required for smooth operation of ICP-MSMS, IC and LA for the first five years should be provided along with the instrument.  a) Suitable aqueous standards (individual/ multielement/internal ) and organometallic standards in suitable organic solvents for ICP-MSMS applications should be provided. |  |
|                | b) Standards, cationic and anionic buffers and mobile phase modifiers required for speciation ofFe, Cr, Hg andAsusing IC should be provided.   |  |
|                | c) Glass standards (610, 612, 614, 616, MACS -3) and other standards (Be,Mg,P,S,V,Ge,Br, Y,Zr,Rh,Pd,Sn,Cs,Ta,W,Pt,Hg,Bi and Pr) for LA applications should be provided.  |  |
| 3              | Software upgrades should be incorporated by the vendor as and when the new versions are released at no additional cost.  |  |
| 4              | Spares and consumables needed for smooth operation of ICP-MSMS, IC and LA for the first five years to be provided along with the instrument.   |  |
| 5              | Accessories such as UPS and Chiller may be supplied locally.   |  |
| 6              | One week onsite training for the operation and general maintenance must be provided.   |  |

(Note: It is mandatory for the bidders to provide the compliance statement in tabular column format along with catalogue page number (comply/not comply) for the above points with document proof as required. Failing which bidders will be technically disqualified)

SIGNATURE OF BIDDER ALONG WITH SEAL OF THE COMPANY WITH DATE

#### FINANCIAL BID (PROFORMA) - BILL OF QUANTITIES (BOQ)

# Item Name: "Triple quad Inductively Coupled Plasma Mass Spectrometer (ICP-MS/MS) with Ion Chromatograph (IC) and Laser Ablation (LA) accessories" Tender No. SAIF/BHAT/90/IOE23/TQICPMSICLA

| It.<br>No | Description of work  | Quantity | Units | Basic<br>Rate<br>in<br>INR | GST in<br>Percentage | Total<br>Amount<br>with taxes<br>in INR |
|-----------|--|----------|-------|----------------------------|----------------------|---|
| 1         | Triple quad Inductively Coupled Plasma Mass<br>Spectrometer (ICP-MS/MS) With Three years<br>standard warranty  | 1        | Nos.  |                            |                      |   |
| 2         | Ion Chromatograph With Three years standard warranty   | 1        | Nos.  |                            |                      |   |
| 3         | Laser ablation system With Three years standard warranty   | 1        | Nos.  |                            |                      |   |
| 4         | Accessories for entire system (ICP-MS/MS + IC+ LA) With Three years standard warranty                          | 1        | Nos.  |                            |                      |   |
| 5         | Additional comprehensive warranty for two years for all parts and peripherals to be quoted separately.         | 1        | Nos.  |                            |                      |   |
| 6         | Suitable auto sampler for IC and LA to be quoted separately.   | 1        | Nos.  |                            |                      |   |
| 7         | Suitable PEEK analytical and guard columns for determination of anions and cations to be quoted separately.    | 1        | Nos.  |                            |                      |   |
| 8         | Suitable PEEK analytical and guard columns for determination of transition metal ions to be quoted separately. | 1        | Nos.  |                            |                      |   |
| 9         | Suitable column oven to be quoted separately.  | 1        | Nos.  |                            |                      |   |
| 10        | Suitable detector for determination of transition metal ions to be quoted separately.                          | 1        | Nos.  |                            |                      |   |
|           | Grand Total  |          |       |                            |                      |   |

| Total Amount Rupees in words |  |
|------------------------------|--|
|------------------------------|--|

Note: Price Quoted for Optional Items(S.No 5 - 10 of this BOQ) will not be considered for price bid evaluation.



# CENTRE FOR INDUSTRIAL CONSULTANCY & SPONSORED RESEARCH (IC&SR) INDIAN INSTITUTE OF TECHNOLOGY MADRAS CHENNAI 600 036



# ELECTRONIC CLEARING SERVICE (Credit Clearing) / REAL TIME GROSS SETTLEMENT (RTGS) FACILITY FOR RECEIVING PAYMENTS A. Details of Account Holder

| Name of the Institution         | Indian Institute of Technology - Madras   |
|---------------------------------|---|
| Complete Contact Address        | Industrial Consultancy and Sponsored Research<br>Indian Institute of Technology-Madras,<br>IIT- Madras Campus Post Office,<br>Sardar Patel Road,<br>Guindy, CHENNAI - 600 036 |
| Telephone No./ Fax No.          | Tel - 044-2257 8356   |
| E- mail ID of the FO/AO/REG/DIR | dricsr@iitm.ac.in   |

#### **B. Bank Account Details:**

| Institution Account Name (As per Bank | The Registrar, Indian Institute of   |
|---------------------------------------|--------------------------------------|
| Record)                               | Technology - Madras                  |
| Account No.                           | 2722101003872                        |
| Account Print Name                    | IIT F A/C , The Registrar IIT Madras |
| IFSC CODE                             | CNRB0002722                          |
| Bank Name (in full)                   | Canara Bank                          |
| Branch Name                           | IIT-Madras Branch                    |
| Complete Branch Address               | Canara Bank,                         |
|                                       | IIT-Madras Branch,                   |
|                                       | IIT- Madras Campus Post Office,      |
|                                       | Sardar Patel Road,                   |
|                                       | Guindy, CHENNAI - 600 036            |
| MICR No.                              | 600015085                            |
| Account Type                          | Savings Account                      |

Certified that the Institute's account is in an RTGS enabled branch.

I hereby declare that the particulars given above are correct and complete.

Date:

Signature of the Competent Authority of the Institution with seal.

# FORMAT FOR AFFIDAVIT OF SELF-CERTIFICATION UNDER PREFERENCE TO MAKE IN INDIA – PER ITEM

| Tender Reference Number:   |  |  |  |
|--|--|--|--|
| Name of the item / Service:  |  |  |  |
| Date:  |  |  |  |
| Hereby solemnly affirm and declare as under:   |  |  |  |
| That I will agree to abide by the terms and conditions of the Public Procurement (Preference to Make in India) Policy vide GoI Order no. P-45021/2/2017-PP (B.EII) dated 15.06.2017 (subsequently revised vide orders dated 28.05.2018, 29.05.2019and 04.06.2020) MOCI order No. 45021/2/2017-PP (BE II) Dt.16th September 2020 & P- 45021/102/2019-BE-II-Part (1) (E-50310) Dt.4th March 2021 and any subsequent modifications/Amendments, if any and |  |  |  |
| That the local content for all inputs which constitute the said item/service/work has been verified by me and I am responsible for the correctness of the claims made therein.   |  |  |  |
| Tick (✓) and Fill the Appropriate Category   |  |  |  |
| I/We [name of the supplier] hereby confirm in respect of quoted items that Local Content is equal to or more than 50% and come under "Class-I Local Supplier" category.  |  |  |  |
| I/We [name of the supplier] hereby confirm in respect of quoted items that Local Content is equal to 20% but less than 50% and come under "Class-II Local Supplier" category.  |  |  |  |
| • The details of the location (s) at which the local value addition is made and the proportionate value of local   |  |  |  |
| content in percentage Address Percentage of Local content:%  |  |  |  |
| For and on behalf of(Name of firm/entity)  |  |  |  |
| Authorized signatory (To be duly authorized by the Board of Directors) <insert and="" contact="" designation="" name,="" no.=""></insert>  |  |  |  |
| [Note: In case of procurement for a value in excess of Rs. 10 Crores, the bidders shall provide this certificate from statutory auditor or cost auditor of the company (in the case of companies) or from a practicing cost accountant or practicing chartered accountant (in respect of suppliers other than companies) giving the percentage of local content.]  |  |  |  |

This letter should be on the letterhead of the quoting firm and should be signed by a competent authority. Non-submission of this will lead to Disqualification of bids.

## Annexure – F

|   | (To be given on the letter head of the bidder)  |
|---|---|
| No  | Dated:  |
|   | CERTIFICATE   |
|   | (Bidders from India)  |
| _   | g restrictions on procurement from a bidder of a country which shares a land ertify that I am not from such a country.  |
|   | OR (whichever is applicable)  |
| (Bidde  | s from Country which shares a land border with India)   |
| border with India and hereby registered with the Competen | g restrictions on procurement from a bidder of a country which shares a land certify that I from (Name of Country) and has been Authority. I also certify that I fulfil all the requirements in this regard and is / evidence of valid registration by the Competent Authority is to be attached) |
| Place:<br>Date:   | Signature of the Tenderer<br>Name & Address of the<br>Tenderer with Office Stamp  |